

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060223\
 Data File : PD075654.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2023 17:36
 Operator : AR\AJ
 Sample : PEM094
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM105

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/02/2023
 Supervised By : Ankita Jodhani 06/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 04:12:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060223CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 02 04:01:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.768	40166863	20641199	20.093	20.359
27)	SA Decachlor...	9.086	7.913	45701950	19006297	19.926	20.239
Target Compounds							
2)	A alpha-BHC	4.003	3.271	31575894	14665789	9.479	9.805
3)	MA gamma-BHC...	4.337	3.601	30325108	13441519	9.699	9.683
6)	B beta-BHC	4.533	3.898	14024030	7071099	9.924	10.297
12)	B 4,4'-DDE	6.207	5.233	376871	357005	0.135m	0.314 #
14)	MA Endrin	6.591	5.635	115.1E6	53264852	45.821	46.396
16)	A 4,4'-DDD	6.723	5.785	913872	628750	0.431m	0.716m#
17)	MA 4,4'-DDT	7.039	6.034	219.8E6	88492806	107.067	108.762
18)	B Endrin al...	6.940	6.108	2604042	1414365	1.259	1.565m
20)	A Methoxychlor	7.515	6.609	305.4E6	120.3E6	244.883	249.179
21)	B Endrin ke...	7.661	6.838	4797475	2463102	1.812	2.106

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060223\
Data File : PD075654.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2023 17:36
Operator : AR\AJ
Sample : PEM094
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 04:12:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060223CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 02 04:01:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
PEM105
Manual Integrations
APPROVED
Reviewed By :Abdul Mirza 06/02/2023
Supervised By :Ankita Jodhani 06/02/2023