

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060320\
 Data File : PD058267.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2020 01:45
 Operator : AJ\MA
 Sample : RESC02
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESC02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 06:09:46 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060320CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 03 05:57:37 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.211	3.867	30738815	42165193	20.409	19.980
2)	SA Decachlor...	7.852	8.861	64088177	82640316	39.181	38.959
Target Compounds							
2)	A alpha-BHC	3.633	4.249	22130216	29612191	9.503	9.348
3)	MA gamma-BHC...	3.904	4.529	20924744	26716863	9.574	9.450
4)	MA Heptachlor	4.184	5.029	13020211	23971645	9.340	8.977
5)	MB Aldrin	4.419	5.330	18256753	25318086	9.229	9.411
6)	B beta-BHC	4.154	4.701	8780007	11420263	9.412	9.365
7)	B delta-BHC	4.345	4.913	19242212	24073712	8.590	8.793
8)	B Heptachlo...	4.852	5.713	16467608	23668987	9.348	9.265
9)	A Endosulfan I	5.179	6.066	16509050	23025178	9.745	9.702
10)	B trans-Chl...	5.072	5.946	16959434	23830247	9.350	9.363
11)	B cis-Chlor...	5.129	6.018	17084468	24000352	9.394	9.396
12)	B 4,4'-DDE	5.298	6.178	34648741	46221364	18.579	18.492
13)	MA Dieldrin	5.417	6.322	34504096	51149425	18.632	19.303
14)	MA Endrin	5.669	6.537	29584455	43312248	19.564	20.171
15)	B Endosulfa...	5.942	6.747	30529999	41151621	18.570	18.682
16)	A 4,4'-DDD	5.806	6.666	22986600	39257429	18.517	18.652
17)	MA 4,4'-DDT	6.041	6.962	24455226	41215411	18.567	19.274
18)	B Endrin al...	6.111	6.870	24736177	36713142	18.765	19.506
19)	B Endosulfa...	6.321	7.093	28651210	42243964	18.007	18.663
20)	A Methoxychlor	6.589	7.422	54238440	111.0E6	95.944	99.527
21)	B Endrin ke...	6.802	7.559	34780874	50789144	19.251	21.083

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060320\
Data File : PD058267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2020 01:45
Operator : AJ\MA
Sample : RESC02
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 06:09:46 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jun 03 05:57:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060320\
Data File : PD058267.D
Acq On : 03 Jun 2020 01:45
Operator : AJ\MA
Sample : RESC02
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RESC02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060320CLP.M
Title : GC Extractables
Last Update : Wed Jun 03 05:57:37 2020
Integrator: ChemStation

RT#1	RT#2	Resolution
3.211	3.633	100.00%
3.633	3.904	100.00%
3.904	4.154	100.00%
4.154	4.184	85.24%
4.184	4.345	100.00%
4.345	4.419	100.00%
4.419	4.852	100.00%
4.852	5.072	100.00%
5.072	5.129	100.00%
5.129	5.179	100.00%
5.179	5.298	100.00%
5.298	5.417	100.00%
5.417	5.509	100.00%
5.509	5.669	100.00%
5.669	5.806	100.00%
5.806	5.942	100.00%
5.942	6.041	98.61%
6.041	6.111	100.00%
6.111	6.321	100.00%
6.321	6.589	100.00%
6.589	6.802	100.00%
6.802	7.852	100.00%

Signal #2


3.867	4.249	100.00%
4.249	4.529	100.00%
4.529	4.701	100.00%
4.701	4.913	100.00%
4.913	5.029	100.00%
5.029	5.330	100.00%
5.330	5.713	100.00%
5.713	5.946	100.00%
5.946	6.018	100.00%
6.018	6.066	97.84%
6.066	6.178	100.00%
6.178	6.322	100.00%
6.322	6.537	100.00%
6.537	6.666	100.00%
6.666	6.747	100.00%
6.747	6.870	100.00%
6.870	6.962	100.00%
6.962	7.093	100.00%
7.093	7.422	100.00%
7.422	7.559	100.00%
7.559	8.861	100.00%

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060320\
Data File : PD058267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2020 01:45
Operator : AJ\MA
Sample : RESC02
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RESC02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 06:02:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jun 03 05:57:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(15) Endosulfan II (B)
5.942min 18.570 ng/ml
response 30529999


(15) Endosulfan II #2 (B)
6.747min 18.682 ng/ml
response 41151621

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060320\
Data File : PD058267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2020 01:45
Operator : AJ\MA
Sample : RESC02
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RESC02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 06:02:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jun 03 05:57:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(6) beta-BHC (B)

4.154min 9.412 ng/ml

response 8780007

(6) beta-BHC #2 (B)

4.701min 9.365 ng/ml


response 11420263

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060320\
Data File : PD058267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2020 01:45
Operator : AJ\MA
Sample : RESC02
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
RESC02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 06:02:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jun 03 05:57:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(11) cis-Chlordane (B)

5.129min 9.394 ng/ml

response 17084468

(11) cis-Chlordane #2 (B)

6.018min 9.396 ng/ml

response 24000352