

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060323\
 Data File : PD075700.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2023 15:27
 Operator : AR\AJ
 Sample : PIBLK156
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 06:01:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060223CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 02 04:01:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.769	45412210	22319939	22.716	22.015
27)	SA Decachlor...	9.088	7.914	114.8E6	46780324	50.064	49.814
Target Compounds							
8)	B Heptachlo...	0.000	4.811f	0	341705	N.D.	0.260
14)	MA Endrin	0.000	5.653	0	365546	N.D.	0.318
15)	B Endosulfa...	0.000	5.887f	0	53407	N.D.	<MDL
20)	A Methoxychlor	0.000	6.599	0	32963	N.D.	<MDL
23)	Toxaphene-2	0.000	5.653	0	365546	N.D.	51.666
25)	Toxaphene-4	0.000	6.599	0	32963	N.D.	1.595


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060323\
Data File : PD075700.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2023 15:27
Operator : AR\AJ
Sample : PIBLK156
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 06:01:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060223CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 02 04:01:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.547 min
Delta R.T.: 0.000 min
Response: 45412210
Conc: 22.72 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.769 min
Delta R.T.: -0.022 min
Response: 22319939
Conc: 22.01 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.697 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 4.811 min
Delta R.T.: 0.066 min
Response: 341705
Conc: 0.26 ng/ml



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 6.590 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#14 Endrin

R.T.: 5.653 min
Delta R.T.: 0.019 min
Response: 365546
Conc: 0.32 ng/ml




#23 Toxaphene-2

R.T.: 0.000 min
Exp R.T.: 6.457 min
Response: 0
Conc: N.D.



#23 Toxaphene-2


R.T.: 5.653 min
Delta R.T.: -0.030 min
Response: 365546
Conc: 51.67 ng/ml



#25 Toxaphene-4


R.T.: 0.000 min
Exp R.T.: 7.166 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#25 Toxaphene-4

R.T.: 6.599 min
Delta R.T.: 0.000 min
Response: 32963
Conc: 1.59 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.088 min
Delta R.T.: 0.002 min
Response: 114825071
Conc: 50.06 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.914 min
Delta R.T.: -0.021 min
Response: 46780324
Conc: 49.81 ng/ml