

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060421\
 Data File : PD063440.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2021 20:20
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 6/4/2021 11:26:13 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 02:32:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 03 01:33:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.971	73455037	679.7E6	53.084	50.728
28)	SA Decachlor...	8.195	9.025	77121171	716.6E6	49.037	51.857
Target Compounds							
2)	A alpha-BHC	3.876	4.359	115.0E6	1098.8E6	54.998	53.158
3)	MA gamma-BHC...	4.159	4.643	106.8E6	1011.3E6	54.388	53.320
4)	MA Heptachlor	4.459	5.151	106.3E6	999.7E6	54.737	52.309
5)	MB Aldrin	4.707	5.455	102.5E6	1020.6E6	54.216	52.490
6)	B beta-BHC	4.410	4.814	44986115	420.0E6	51.711	53.344
7)	B delta-BHC	4.614	5.031	108.3E6	956.8E6	55.215	50.999
8)	B Heptachlo...	5.150	5.839	93740730	915.1E6	53.494	50.058m
9)	A Endosulfan I	5.489	6.197	87399300	873.0E6	53.025	51.237
10)	B gamma-Chl...	5.377	6.075	95329447	966.9E6	53.380	52.047
11)	B alpha-Chl...	5.436	6.147	93460812	912.0E6	52.854	51.299
12)	B 4,4'-DDE	5.599	6.305	94009216	915.4E6	54.067	50.372
13)	MA Dieldrin	5.732	6.454	93602164	922.3E6	53.292	50.951
14)	MA Endrin	5.990	6.672	85391445	788.1E6	53.690	51.271
15)	B Endosulfa...	6.263	6.881	76548546	739.8E6	54.597	50.172
16)	A 4,4'-DDD	6.115	6.794	77849331	753.1E6	53.680	50.613
17)	MA 4,4'-DDT	6.357	7.093	77809983	765.0E6	53.858	50.821
18)	B Endrin al...	6.432	7.005	60573653	594.5E6	50.909	49.069
19)	B Endosulfa...	6.645	7.228	72607776	746.1E6	56.813	49.203
20)	A Methoxychlor	6.903	7.551	41228695	371.1E6	51.693	50.101
21)	B Endrin ke...	7.138	7.698	92082701	742.2E6	51.904	50.759
22)	Mirex	7.325	8.151	65136909	515.4E6	56.742	51.548

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060421\
Data File : PD063440.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2021 20:20
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
6/4/2021 11:26:13 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 02:32:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
Quant Title : GC Extractables
QLast Update : Thu Jun 03 01:33:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

