

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060422\
 Data File : PD070040.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2022 20:10
 Operator : AR\AJ
 Sample : N2990-21MS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 P028-SS001-0612-01MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/04/2022
 Supervised By : Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 00:43:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052622.M
 Quant Title : GC Extractables
 QLast Update : Thu May 26 05:06:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.442	4.170	17089647	160.6E6	16.059	18.226
28)	SA Decachlor...	8.195	9.365	14651534	81668595	13.433	15.101
Target Compounds							
2)	A alpha-BHC	3.881	4.566	95168997	824.0E6	58.075	62.924
3)	MA gamma-BHC...	4.164	4.857	89792731	788.0E6	52.026	63.879
4)	MA Heptachlor	4.461	5.384	85953676	747.2E6	57.574	64.187
5)	MB Aldrin	4.709	5.695	77691233	748.8E6	53.608	64.909m
6)	B beta-BHC	4.415	5.019	37138183	323.6E6	53.361	61.600m
7)	B delta-BHC	4.618	5.247	86413182	623.1E6	56.717	65.498
8)	B Heptachlo...	5.152	6.085	77335236	567.8E6	57.912	72.164
9)	A Endosulfan I	5.491	6.447	70937853	574.1E6	56.615	62.950
10)	B gamma-Chl...	5.378	6.322	77554635	640.1E6	57.314	64.587
11)	B alpha-Chl...	5.436	6.396	75766817	612.3E6	56.911	63.001
12)	B 4,4'-DDE	5.602	6.546	70946919	576.0E6	55.998	63.660
13)	MA Dieldrin	5.734	6.706	77119387	608.2E6	58.587	64.464
14)	MA Endrin	5.992	6.926	70467330	493.4E6	61.028	73.104
15)	B Endosulfa...	6.267	7.132	66529714	484.0E6	59.811	65.160
16)	A 4,4'-DDD	6.119	7.038	61788102	485.3E6	57.106	65.528
17)	MA 4,4'-DDT	6.359	7.340	63737763	482.5E6	64.312	65.936
18)	B Endrin al...	6.435	7.255	51841141	373.1E6	60.337	64.988
19)	B Endosulfa...	6.648	7.480	63669883	455.6E6	55.234	61.791
20)	A Methoxychlor	6.909	7.794	40969546	229.2E6	72.218	64.285
21)	B Endrin ke...	7.142	7.956	77707609	470.4E6	58.314	59.084
22)	Mirex	7.324	8.420	57837726	323.7E6	57.520	60.593m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060422\
Data File : PD070040.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2022 20:10
Operator : AR\AJ
Sample : N2990-21MS
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
P028-SS001-0612-01MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/04/2022
Supervised By :Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 00:43:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052622.M
Quant Title : GC Extractables
QLast Update : Thu May 26 05:06:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

