

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060422\
 Data File : PD070045.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2022 21:21
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/04/2022
 Supervised By : Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 00:45:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052622.M
 Quant Title : GC Extractables
 QLast Update : Thu May 26 05:06:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.442	4.169	51664052	459.6E6	48.547	52.154
28)	SA Decachlor...	8.194	9.364	52497527	295.7E6	48.130	54.674
Target Compounds							
2)	A alpha-BHC	3.881	4.565	80001127	699.7E6	48.819	53.430
3)	MA gamma-BHC...	4.164	4.855	76494735	648.2E6	44.321	52.546
4)	MA Heptachlor	4.461	5.382	72737151	624.4E6	48.722	53.635
5)	MB Aldrin	4.708	5.695	70315758	607.6E6	48.519	52.672
6)	B beta-BHC	4.415	5.019	33136910	293.9E6	47.612	55.944
7)	B delta-BHC	4.618	5.245	75277419	517.7E6	49.408	54.414
8)	B Heptachlo...	5.152	6.083	64733933	468.9E6	48.476	59.587
9)	A Endosulfan I	5.490	6.446	61074605	489.2E6	48.743	53.639
10)	B gamma-Chl...	5.377	6.320	65513960	531.6E6	48.416	53.639
11)	B alpha-Chl...	5.436	6.394	64644487	517.9E6	48.557	53.292
12)	B 4,4'-DDE	5.602	6.545	61431279	484.4E6	48.487	53.540
13)	MA Dieldrin	5.734	6.705	64096075	505.8E6	48.694	53.614
14)	MA Endrin	5.992	6.924	56347550	363.8E6	48.800	53.896
15)	B Endosulfa...	6.265	7.130	53431865	392.0E6	48.036	52.777
16)	A 4,4'-DDD	6.119	7.036	52945480	402.2E6	48.933	54.318
17)	MA 4,4'-DDT	6.359	7.339	49389528	389.9E6	49.835	53.279
18)	B Endrin al...	6.435	7.254	41621064	305.5E6	48.442	53.198
19)	B Endosulfa...	6.648	7.479	54313860	391.1E6	47.118	53.042
20)	A Methoxychlor	6.909	7.793	29142046	196.0E6	51.370	54.981
21)	B Endrin ke...	7.141	7.954	64392824	408.8E6	48.323	51.344
22)	Mirex	7.323	8.418	49128376	289.8E6	48.859	54.256m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060422\
Data File : PD070045.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2022 21:21
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/04/2022
Supervised By : Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 00:45:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052622.M
Quant Title : GC Extractables
QLast Update : Thu May 26 05:06:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

