

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060521\
 Data File : PD063463.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jun 2021 10:42
 Operator : AR\AJ
 Sample : M2585-02DL 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ETGI-363DL

Manual Integrations
 APPROVED

Ankita
 6/7/2021 8:41:56 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 05 02:40:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 03 01:33:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.971	2028273	21308634	1.466	1.590
28)	SA Decachlor...	8.196	9.026	2636914	25281409	1.677	1.830
Target Compounds							
10)	B gamma-Chl...	5.377	6.074	34619294	816.8E6	19.385m	43.971 #
11)	B alpha-Chl...	5.434	6.149	87653419	1712.4E6	49.570m	96.318 #
12)	B 4,4'-DDE	5.600	6.302	1875577	40661264	1.079m	2.237 #
14)	MA Endrin	5.988	6.665	2144899	18734549	1.349m	1.219m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060521\
Data File : PD063463.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jun 2021 10:42
Operator : AR\AJ
Sample : M2585-02DL 10X
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
ETGI-363DL


Manual Integrations
APPROVED

Ankita
6/7/2021 8:41:56 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 05 02:40:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
Quant Title : GC Extractables
QLast Update : Thu Jun 03 01:33:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.440 min
Delta R.T.: 0.000 min
Response: 2028273
Conc: 1.47 ng/ml

Instrument :
ECD_D
ClientSampleId :
ETGI-363DL


Manual Integrations
APPROVED

Ankita
6/7/2021 8:41:56 AM




#1 Tetrachloro-m-xylene

R.T.: 3.971 min
Delta R.T.: 0.000 min
Response: 21308634
Conc: 1.59 ng/ml




#10 gamma-Chlordane

R.T.: 5.377 min
Delta R.T.: 0.000 min
Response: 34619294
Conc: 19.39 ng/ml m



#10 gamma-Chlordane

R.T.: 6.074 min
Delta R.T.: 0.000 min
Response: 816826361
Conc: 43.97 ng/ml




#11 alpha-Chlordane

R.T.: 5.434 min
Delta R.T.: -0.001 min
Response: 87653419
Conc: 49.57 ng/ml m

Instrument :
ECD_D
ClientSampleId :
ETGI-363DL


Manual Integrations
APPROVED

Ankita
6/7/2021 8:41:56 AM




#11 alpha-Chlordane

R.T.: 6.149 min
Delta R.T.: 0.002 min
Response: 1712370337
Conc: 96.32 ng/ml




#12 4,4'-DDE

R.T.: 5.600 min
Delta R.T.: 0.001 min
Response: 1875577
Conc: 1.08 ng/ml m



#12 4,4'-DDE

R.T.: 6.302 min
Delta R.T.: -0.002 min
Response: 40661264
Conc: 2.24 ng/ml




#14 Endrin

R.T.: 5.988 min
Delta R.T.: -0.001 min
Response: 2144899
Conc: 1.35 ng/ml m

Instrument :
ECD_D
ClientSampleId :
ETGI-363DL


Manual Integrations
APPROVED

Ankita
6/7/2021 8:41:56 AM




#14 Endrin

R.T.: 6.665 min
Delta R.T.: -0.006 min
Response: 18734549
Conc: 1.22 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.196 min
Delta R.T.: 0.002 min
Response: 2636914
Conc: 1.68 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.026 min
Delta R.T.: 0.002 min
Response: 25281409
Conc: 1.83 ng/ml