

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060522\
 Data File : PD070050.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jun 2022 09:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/06/2022
 Supervised By : Ankita Jodhani 06/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 10:02:51 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052622.M
 Quant Title : GC Extractables
 QLast Update : Thu May 26 05:06:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.442	4.168	53948284	516.5E6	50.693	58.611
28)	SA Decachlor...	8.195	9.362	52181390	302.6E6	47.840	55.961
Target Compounds							
2)	A alpha-BHC	3.880	4.565	82228078	775.9E6	50.178	59.253
3)	MA gamma-BHC...	4.164	4.855	79176184	715.5E6	45.874	58.004 #
4)	MA Heptachlor	4.461	5.381	74284913	688.2E6	49.758	59.112
5)	MB Aldrin	4.708	5.694	72243544	663.5E6	49.849	57.513
6)	B beta-BHC	4.415	5.019	33849287	320.5E6	48.635	61.012 #
7)	B delta-BHC	4.618	5.243	76703541	659.4E6	50.344	69.310m#
8)	B Heptachlo...	5.151	6.082	66865814	587.3E6	50.072	74.635 #
9)	A Endosulfan I	5.490	6.445	62540640	531.0E6	49.913	58.224
10)	B gamma-Chl...	5.377	6.319	67310800	579.3E6	49.744	58.454
11)	B alpha-Chl...	5.436	6.394	66407585	561.7E6	49.881	57.796
12)	B 4,4'-DDE	5.601	6.544	63542106	526.1E6	50.153	58.148
13)	MA Dieldrin	5.733	6.703	65832266	549.8E6	50.012	58.274
14)	MA Endrin	5.992	6.921	57060994	375.4E6	49.418	55.621m
15)	B Endosulfa...	6.266	7.129	55058548	421.5E6	49.498	56.752
16)	A 4,4'-DDD	6.118	7.035	54451487	434.5E6	50.325	58.674
17)	MA 4,4'-DDT	6.358	7.338	50520677	417.5E6	50.976	57.051
18)	B Endrin al...	6.435	7.253	43436031	332.1E6	50.554	57.845
19)	B Endosulfa...	6.648	7.478	56895127	413.9E6	49.357	56.131
20)	A Methoxychlor	6.909	7.791	29759796	207.3E6	52.459	58.137
21)	B Endrin ke...	7.141	7.952	66334597	438.0E6	49.780	55.018
22)	Mirex	7.324	8.419	50767353	298.3E6	50.489	55.841

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060522\
Data File : PD070050.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jun 2022 09:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/06/2022
Supervised By :Ankita Jodhani 06/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 10:02:51 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD052622.M
Quant Title : GC Extractables
QLast Update : Thu May 26 05:06:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

