

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060525\
 Data File : PD088802.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2025 14:08
 Operator : AR\AJ
 Sample : TOXAPH101
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 TOXAPH1035

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 05 16:55:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060525CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 05 16:40:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.880	14449595	92122603	5.155	5.444
27)	SA Decachlor...	9.071	8.072	35807893	177.9E6	11.003	11.269
Target Compounds							
22)	Toxaphene-1	6.240	5.147	14958658	63806290	525.686	554.992
23)	Toxaphene-2	6.440	6.119	20266262	70768693	507.137	562.762
24)	Toxaphene-3	7.055	6.756	52416376	211.4E6	484.332	520.046
25)	Toxaphene-4	7.147	7.198	38680376	150.3E6	482.254	543.718
26)	Toxaphene-5	7.928	7.329	29004692	106.7E6	494.617	502.553

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060525\
Data File : PD088802.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2025 14:08
Operator : AR\AJ
Sample : TOXAPH101
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 05 16:55:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060525CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 05 16:40:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

