

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060621\
 Data File : PD063495.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2021 10:17
 Operator : AR\AJ
 Sample : M2589-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 18-B7(2-4)

Manual Integrations
 APPROVED

Ankita
 6/7/2021 4:08:41 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 08:14:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 03 01:33:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.439	3.971	11768420	123.9E6	8.505	9.246
28)	SA Decachlor...	8.195	9.026	13765232	133.5E6	8.753	9.664
Target Compounds							
4)	MA Heptachlor	4.460	5.151	3734516	24325553	1.923	1.273 #
10)	B gamma-Chl...	5.377	6.075	7090448	128.0E6	3.970m	6.889 #
11)	B alpha-Chl...	5.435	6.149	10582733	242.5E6	5.985	13.638 #
12)	B 4,4'-DDE	5.598	6.302	5636199	69065570	3.242m	3.800
14)	MA Endrin	5.988	6.664	2024496	10836079	1.273m	0.705 #
17)	MA 4,4'-DDT	6.354	7.092	7144706	21605994	4.945m	1.435 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060621\
Data File : PD063495.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2021 10:17
Operator : AR\AJ
Sample : M2589-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
18-B7(2-4)

Manual Integrations
APPROVED

Ankita
6/7/2021 4:08:41 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 08:14:53 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
Quant Title : GC Extractables
QLast Update : Thu Jun 03 01:33:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

