

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060621\
 Data File : PD063501.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2021 11:41
 Operator : AR\AJ
 Sample : M2589-09
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 18-B9(4-6)

Manual Integrations
 APPROVED

Ankita
 6/7/2021 4:08:51 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 08:15:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 03 01:33:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.971	16004063	149.0E6	11.566	11.118
28)	SA Decachlor...	8.196	9.026	14999793	129.2E6	9.538	9.348
Target Compounds							
10)	B gamma-Chl...	5.377	6.078	7573771	177.9E6	4.241m	9.577 #
11)	B alpha-Chl...	5.436	6.148	12927713	239.8E6	7.311	13.486 #
12)	B 4,4'-DDE	5.600	6.304	98325643	815.9E6	56.550m	44.898
14)	MA Endrin	5.987	6.664	6280085	214.3E6	3.949m	13.943 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060621\
Data File : PD063501.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2021 11:41
Operator : AR\AJ
Sample : M2589-09
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
18-B9(4-6)

Manual Integrations
APPROVED

Ankita
6/7/2021 4:08:51 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 08:15:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
Quant Title : GC Extractables
QLast Update : Thu Jun 03 01:33:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

