

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060621\
 Data File : PD063507.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2021 13:04
 Operator : AR\AJ
 Sample : M2589-15
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 18-B11(2-4)

Manual Integrations
 APPROVED

Ankita
 6/7/2021 4:10:08 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 08:16:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 03 01:33:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.970	15738230	129.9E6	11.374	9.696m
28)	SA Decachlor...	8.196	9.026	15316116	134.2E6	9.739	9.714
Target Compounds							
5)	MB Aldrin	4.708	5.453	4976422	23543374	2.633m	1.211 #
10)	B gamma-Chl...	5.376	6.075	17773932	173.7E6	9.953	9.348
11)	B alpha-Chl...	5.436	6.149	19028782	288.9E6	10.761	16.249 #
12)	B 4,4'-DDE	5.599	6.303	38473098	321.0E6	22.127m	17.666
13)	MA Dieldrin	5.725	6.445	4117063	21996345	2.344m	1.215 #
14)	MA Endrin	5.993	6.664	6439711	24915100	4.049m	1.621 #
16)	A 4,4'-DDD	6.115	6.797	4706623	110.6E6	3.245m	7.430m#
17)	MA 4,4'-DDT	6.354	7.092	13285884	33354305	9.196m	2.216m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060621\
Data File : PD063507.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2021 13:04
Operator : AR\AJ
Sample : M2589-15
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
18-B11(2-4)

Manual Integrations
APPROVED

Ankita
6/7/2021 4:10:08 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 08:16:48 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
Quant Title : GC Extractables
QLast Update : Thu Jun 03 01:33:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

