

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060621\
 Data File : PD063520.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2021 16:35
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 08:18:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 03 01:33:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.972	72613844	692.7E6	52.476	51.696
28)	SA Decachlor...	8.197	9.027	55383826	441.1E6	35.216	31.923
Target Compounds							
2)	A alpha-BHC	3.876	4.360	112.4E6	1064.3E6	53.740	51.487
3)	MA gamma-BHC...	4.159	4.644	104.3E6	970.7E6	53.101	51.178
4)	MA Heptachlor	4.460	5.152	100.6E6	939.0E6	51.806	49.137
5)	MB Aldrin	4.708	5.456	97312389	939.3E6	51.491	48.306
6)	B beta-BHC	4.410	4.815	43810805	396.7E6	50.360	50.387
7)	B delta-BHC	4.614	5.031	102.1E6	940.1E6	52.061	50.050
8)	B Heptachlo...	5.151	5.842	89460683	809.1E6	51.052	44.259
9)	A Endosulfan I	5.490	6.198	82048022	720.7E6	49.779	42.300
10)	B gamma-Chl...	5.378	6.075	89843837	800.0E6	50.309	43.063
11)	B alpha-Chl...	5.437	6.148	88270539	758.8E6	49.919	42.679
12)	B 4,4'-DDE	5.601	6.305	87383779	745.1E6	50.257	41.000
13)	MA Dieldrin	5.733	6.454	86073652	731.6E6	49.005	40.417
14)	MA Endrin	5.991	6.672	73893563	568.9E6	46.461	37.011
15)	B Endosulfa...	6.264	6.882	65594838	551.6E6	46.784	37.407
16)	A 4,4'-DDD	6.117	6.795	68392670	553.1E6	47.160	37.175
17)	MA 4,4'-DDT	6.358	7.093	61731424	499.1E6	42.729	33.159
18)	B Endrin al...	6.433	7.005	50455665	424.4E6	42.405	35.031
19)	B Endosulfa...	6.647	7.229	61028412	513.9E6	47.753	33.887 #
20)	A Methoxychlor	6.905	7.552	31791851	240.8E6	39.861	32.517
21)	B Endrin ke...	7.138	7.699	70566853	496.6E6	39.777	33.960
22)	Mirex	7.327	8.153	49431771	321.0E6	43.061	32.105 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060621\
Data File : PD063520.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2021 16:35
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 08:18:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060221.M
Quant Title : GC Extractables
QLast Update : Thu Jun 03 01:33:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

