

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060624\
 Data File : PD083441.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2024 20:03
 Operator : AR\AJ
 Sample : PIBLK167
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BHBY2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 05 22:01:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060424CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 04 07:04:40 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.564	2.773	27528750	99857708	19.189	21.082
2)	SA Decachlor...	9.119	7.900	35985048	101.1E6	17.610	20.277
Target Compounds							
2)	A alpha-BHC	0.000	3.302f	0	588086	N.D.	<MDL
3)	MA gamma-BHC...	0.000	3.577	0	284772	N.D.	<MDL
4)	MA Heptachlor	0.000	3.924	0	62871	N.D.	<MDL
5)	MB Aldrin	0.000	4.200	0	12145318	N.D.	1.876
7)	B delta-BHC	0.000	4.125	0	7485496	N.D.	1.062
8)	B Heptachlo...	5.713	4.742	2104402	1458712	0.958	0.245 #
9)	A Endosulfan I	0.000	5.132f	0	4220794	N.D.	0.774
10)	B trans-Chl...	0.000	4.965	0	2019108	N.D.	0.345
11)	B cis-Chlor...	0.000	5.040	0	7315992	N.D.	1.227
13)	MA Dieldrin	0.000	5.335	0	1749187	N.D.	0.288
14)	MA Endrin	0.000	5.647	0	1461374	N.D.	0.273
16)	A 4,4'-DDD	0.000	5.801	0	39855	N.D.	<MDL
19)	B Endosulfa...	0.000	6.352	0	432231	N.D.	<MDL
22)	Toxaphene-1	5.832f	5.010	3671269	2234400	555.386	75.476 #
23)	Toxaphene-2	0.000	5.647	0	1461374	N.D.	43.853
26)	Toxaphene-5	0.000	6.981f	0	589686	N.D.	10.708


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060624\
Data File : PD083441.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2024 20:03
Operator : AR\AJ
Sample : PIBLK167
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 05 22:01:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060424CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 04 07:04:40 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.564 min
Delta R.T.: 0.002 min
Response: 27528750
Conc: 19.19 ng/ml

Instrument :
ECD_D
ClientSampleId :
BHY2




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: 0.001 min
Response: 99857708
Conc: 21.08 ng/ml




#5 Aldrin

R.T.: 0.000 min
Exp R.T. : 5.289 min
Response: 0
Conc: N.D.



#5 Aldrin


R.T.: 4.200 min
Delta R.T.: -0.018 min
Response: 12145318
Conc: 1.88 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T.: 4.800 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BHBY2




#7 delta-BHC

R.T.: 4.125 min
Delta R.T.: -0.002 min
Response: 7485496
Conc: 1.06 ng/ml




#8 Heptachlor epoxide

R.T.: 5.713 min
Delta R.T.: -0.004 min
Response: 2104402
Conc: 0.96 ng/ml



#8 Heptachlor epoxide


R.T.: 4.742 min
Delta R.T.: 0.023 min
Response: 1458712
Conc: 0.24 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 6.105 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BHBV2




#9 Endosulfan I

R.T.: 5.132 min
Delta R.T.: 0.044 min
Response: 4220794
Conc: 0.77 ng/ml




#10 trans-Chlordane

R.T.: 0.000 min
Exp R.T.: 5.974 min
Response: 0
Conc: N.D.



#10 trans-Chlordane


R.T.: 4.965 min
Delta R.T.: -0.004 min
Response: 2019108
Conc: 0.35 ng/ml



#11 cis-Chlordane


R.T.: 0.000 min
Exp R.T.: 6.053 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BHY2




#11 cis-Chlordane

R.T.: 5.040 min
Delta R.T.: 0.008 min
Response: 7315992
Conc: 1.23 ng/ml




#13 Dieldrin

R.T.: 0.000 min
Exp R.T.: 6.381 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 5.335 min
Delta R.T.: -0.017 min
Response: 1749187
Conc: 0.29 ng/ml



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 6.612 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BHY2




#14 Endrin

R.T.: 5.647 min
Delta R.T.: 0.020 min
Response: 1461374
Conc: 0.27 ng/ml




#22 Toxaphene-1

R.T.: 5.832 min
Delta R.T.: -0.052 min
Response: 3671269
Conc: 555.39 ng/ml



#22 Toxaphene-1


R.T.: 5.010 min
Delta R.T.: 0.017 min
Response: 2234400
Conc: 75.48 ng/ml



#23 Toxaphene-2


R.T.: 0.000 min
Exp R.T.: 6.477 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BHY2




#23 Toxaphene-2

R.T.: 5.647 min
Delta R.T.: -0.029 min
Response: 1461374
Conc: 43.85 ng/ml




#26 Toxaphene-5

R.T.: 0.000 min
Exp R.T.: 7.611 min
Response: 0
Conc: N.D.



#26 Toxaphene-5


R.T.: 6.981 min
Delta R.T.: -0.049 min
Response: 589686
Conc: 10.71 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.119 min
Delta R.T.: 0.001 min
Response: 35985048
Conc: 17.61 ng/ml

Instrument :
ECD_D
ClientSampleId :
BHY2



#27 Decachlorobiphenyl

R.T.: 7.900 min
Delta R.T.: 0.000 min
Response: 101127919
Conc: 20.28 ng/ml