

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(9) Endosulfan I (A)

6.094min 1.921 ng/ml

response 3756025

(9) Endosulfan I #2 (A)

5.118min 19.848 ng/ml

response 131209988

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(9) Endosulfan I (A)
6.138min 15.859 ng/ml m
response 31012772

(9) Endosulfan I #2 (A)
5.117min 14.611 ng/ml m
response 96588279

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(13) Dieldrin (MA)
6.358min 2.633 ng/ml
response 5444202

(13) Dieldrin #2 (MA)
5.362min 12.708 ng/ml
response 92581818

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)
6.403min 8.026 ng/ml m
response 16595491

(13) Dieldrin #2 (MA)
5.361min 9.567 ng/ml m
response 69699987

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(16) 4,4'-DDD (A)

6.733min 9.875 ng/ml

response 14469947

(16) 4,4'-DDD #2 (A)

5.779min 2.595 ng/ml

response 14628690

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(16) 4,4'-DDD (A)

6.731min 10.777 ng/ml m

response 15791855

(16) 4,4'-DDD #2 (A)

5.777min 5.993 ng/ml m

response 33788050

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.063min 3.361 ng/ml

response 5062959

(17) 4,4'-DDT #2 (MA)

6.024min 1.677 ng/ml

response 9495042

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.061min 2.723 ng/ml m

response 4102538

(17) 4,4'-DDT #2 (MA)

6.022min 3.275 ng/ml m

response 18540667

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.113min 9.507 ng/ml

response 18485696

(27) Decachlorobiphenyl #2 (SA)

7.900min 16.372 ng/ml

response 99414152

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060724\
Data File : PD083480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 23:52
Operator : AR\AJ
Sample : P2590-06
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BHBL2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 12:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.111min 14.035 ng/ml m

response 27291311

(27) Decachlorobiphenyl #2 (SA)

7.900min 16.372 ng/ml

response 99414152