

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060822\
 Data File : PD070225.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jun 2022 20:19
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 23:41:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060622.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 08 02:48:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.441	4.169	56510005	510.9E6	52.052	52.343
28)	SA Decachlor...	8.192	9.362	53881921	307.3E6	48.450	49.968
Target Compounds							
2)	A alpha-BHC	3.879	4.565	86402351	782.4E6	52.826	53.687
3)	MA gamma-BHC...	4.162	4.855	79732624	715.8E6	51.780	53.044
4)	MA Heptachlor	4.459	5.382	72988361	654.4E6	47.200	49.990
5)	MB Aldrin	4.706	5.695	75500697	684.0E6	51.337	53.475
6)	B beta-BHC	4.413	5.020	35808218	321.9E6	49.773	52.873
7)	B delta-BHC	4.617	5.245	81599468	683.6E6	52.710	75.305 #
8)	B Heptachlo...	5.150	6.084	69576351	609.9E6	50.739	74.033 #
9)	A Endosulfan I	5.488	6.446	65084979	550.0E6	49.723	51.932
10)	B gamma-Chl...	5.376	6.321	70342749	616.1E6	50.624	53.605
11)	B alpha-Chl...	5.434	6.395	69201769	586.0E6	49.978	51.840
12)	B 4,4'-DDE	5.600	6.545	64280151	538.9E6	51.399	51.401
13)	MA Dieldrin	5.732	6.705	68521309	561.2E6	50.878	51.344
14)	MA Endrin	5.990	6.925	56231327	422.9E6	46.693	50.378
15)	B Endosulfa...	6.263	7.130	57219430	439.1E6	50.680	51.932
16)	A 4,4'-DDD	6.117	7.037	56374159	455.0E6	51.906	54.099
17)	MA 4,4'-DDT	6.357	7.339	49790967	399.8E6	46.933	46.581
18)	B Endrin al...	6.433	7.254	45862760	340.9E6	49.558	49.680
19)	B Endosulfa...	6.646	7.479	58974444	426.6E6	49.829	51.219
20)	A Methoxychlor	6.907	7.793	28217391	198.1E6	45.768	46.987
21)	B Endrin ke...	7.140	7.954	69110162	455.4E6	51.342	52.479
22)	Mirex	7.321	8.420	50336833	306.4E6	48.199	49.885

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD060822\
Data File : PD070225.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jun 2022 20:19
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 23:41:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060622.M
Quant Title : GC Extractables
QLast Update : Wed Jun 08 02:48:31 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

