

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061025\
 Data File : PD088885.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2025 19:22
 Operator : AR\AJ
 Sample : Q2266-01MS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 WC-3MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 07:43:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
 Quant Title : GC Extractables
 QLast Update : Mon May 19 15:27:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.548	2.880	40789417	269.5E6	18.851	17.814
2) SA Decachlor...	9.072	8.072	56161470	268.4E6	16.410	14.704
Target Compounds						
2) A alpha-BHC	3.998	3.392	245.5E6	1152.4E6	51.388	48.212
3) MA gamma-BHC...	4.329	3.729	235.5E6	1064.6E6	51.162	48.027
4) MA Heptachlor	4.928	4.082	216.9E6	1020.5E6	48.506	45.448
5) MB Aldrin	5.270	4.368	225.4E6	1054.7E6	51.285	48.095
6) B beta-BHC	4.514	4.025	91670510	469.2E6	50.839	48.163
7) B delta-BHC	4.763	4.261	231.8E6	1083.3E6	54.842	48.591
8) B Heptachlo...	5.690	4.872	199.6E6	939.7E6	50.277	47.287
9) A Endosulfan I	6.073	5.246	187.4E6	900.8E6	49.815	47.464
10) B gamma-Chl...	5.945	5.125	197.5E6	1022.9E6	49.508	47.950
11) B alpha-Chl...	6.026	5.189	199.7E6	989.5E6	49.841	47.977
12) B 4,4'-DDE	6.195	5.375	181.0E6	979.4E6	50.426	46.857
13) MA Dieldrin	6.346	5.512	198.2E6	978.4E6	49.394	46.495
14) MA Endrin	6.573	5.788	166.0E6	885.4E6	48.593	45.911
15) B Endosulfa...	6.785	6.080	163.9E6	840.0E6	47.589	45.845
16) A 4,4'-DDD	6.704	5.929	141.4E6	787.1E6	50.760	45.258
17) MA 4,4'-DDT	7.020	6.182	142.2E6	792.9E6	45.534	43.718
18) B Endrin al...	6.914	6.258	123.0E6	623.2E6	47.793	44.744
19) B Endosulfa...	7.148	6.481	150.7E6	799.9E6	47.058	45.051
20) A Methoxychlor	7.492	6.753	72201418	406.5E6	43.220	42.460
21) B Endrin ke...	7.629	6.990	161.9E6	869.6E6	47.297	44.946
22) Mirex	8.113	7.185	116.9E6	635.8E6	44.982	41.820

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061025\
Data File : PD088885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2025 19:22
Operator : AR\AJ
Sample : Q2266-01MS
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
WC-3MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 07:43:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051925.M
Quant Title : GC Extractables
QLast Update : Mon May 19 15:27:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

