

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061223\
 Data File : PD075769.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2023 17:08
 Operator : AR\AJ
 Sample : PIBLK164
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 22:33:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060223CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 02 04:01:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.769	46340056	21951967	23.181	21.652
27)	SA Decachlor...	9.093	7.916	125.4E6	47402558	54.662	50.476
Target Compounds							
4)	MA Heptachlor	4.944	0.000	2559941	0	0.789	N.D. #
26)	Toxaphene-5	0.000	6.757	0	91897	N.D.	2.838


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061223\
Data File : PD075769.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2023 17:08
Operator : AR\AJ
Sample : PIBLK164
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 22:33:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060223CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 02 04:01:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.547 min
Delta R.T.: 0.002 min
Response: 46340056
Conc: 23.18 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.769 min
Delta R.T.: -0.022 min
Response: 21951967
Conc: 21.65 ng/ml




#4 Heptachlor

R.T.: 4.944 min
Delta R.T.: 0.017 min
Response: 2559941
Conc: 0.79 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T. : 3.941 min
Response: 0
Conc: N.D.



#26 Toxaphene-5


R.T.: 0.000 min
Exp R.T.: 7.589 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#26 Toxaphene-5

R.T.: 6.757 min
Delta R.T.: 0.031 min
Response: 91897
Conc: 2.84 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.093 min
Delta R.T.: 0.008 min
Response: 125370057
Conc: 54.66 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.916 min
Delta R.T.: -0.019 min
Response: 47402558
Conc: 50.48 ng/ml