

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061224\
 Data File : PD083673.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2024 13:12
 Operator : AR\AJ
 Sample : P2796-07
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC-BLANK-SPIKE

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/14/2024
 Supervised By :Ankita Jodhani 06/14/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 01:26:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 07 04:03:14 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds


Target Compounds						
3) MA	gamma-BHC...	4.363	3.597	39743392	152.3E6	18.268 19.505m
4) MA	Heptachlor	4.956	3.935	42898521	151.2E6	17.659 19.312m
5) MB	Aldrin	5.299	4.214	41876798	154.5E6	17.627 20.000m
13) MA	Dieldrin	6.389	5.349	82292933	292.5E6	39.799m 40.145m
14) MA	Endrin	6.621	5.624	73589591	257.5E6	43.563m 39.817m
17) MA	4,4'-DDT	7.070	6.022	60320261	216.7E6	40.041m 38.275m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061224\
Data File : PD083673.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2024 13:12
Operator : AR\AJ
Sample : P2796-07
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 01:26:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Mirza 06/14/2024
Supervised By: Ankita Jodhani 06/14/2024

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sample :
PEST-GPC-BLANK-SPIKE