

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061322\
 Data File : PD070264.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2022 17:26
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 04:27:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061322.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 14 04:26:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.402	4.123	107.2E6	873.6E6	76.399	75.860
2)	SA Decachlor...	8.134	9.281	106.5E6	500.5E6	76.123	75.417
Target Compounds							
2)	A alpha-BHC	3.838	4.517	175.2E6	1359.8E6	76.794	75.950
3)	MA gamma-BHC...	4.118	4.805	160.5E6	1225.3E6	76.673	75.844
4)	MA Heptachlor	4.412	5.328	156.9E6	1163.0E6	76.587	76.025
5)	MB Aldrin	4.657	5.639	152.6E6	1113.8E6	76.678	76.068
6)	B beta-BHC	4.368	4.969	67012596	524.6E6	76.083	75.607
7)	B delta-BHC	4.569	5.193	165.0E6	849.4E6	76.636	76.753
8)	B Heptachlo...	5.098	6.026	140.0E6	744.0E6	76.671	76.381
9)	A Endosulfan I	5.435	6.387	130.1E6	889.8E6	76.510	75.641
10)	B gamma-Chl...	5.323	6.262	142.3E6	978.1E6	76.599	75.588
11)	B alpha-Chl...	5.381	6.336	139.0E6	951.6E6	76.604	75.672
12)	B 4,4'-DDE	5.548	6.487	135.7E6	932.4E6	76.729	75.827
13)	MA Dieldrin	5.677	6.644	142.0E6	942.9E6	76.622	75.800
14)	MA Endrin	5.934	6.864	129.6E6	785.9E6	76.464	75.591
15)	B Endosulfa...	6.207	7.069	120.0E6	732.5E6	76.293	75.663
16)	A 4,4'-DDD	6.063	6.978	117.1E6	756.9E6	76.582	75.656
17)	MA 4,4'-DDT	6.302	7.280	115.1E6	780.2E6	76.960	75.949
18)	B Endrin al...	6.377	7.193	90016909	556.5E6	76.429	75.634
19)	B Endosulfa...	6.589	7.418	120.1E6	719.0E6	76.425	75.795
20)	A Methoxychlor	6.852	7.734	61540383	368.0E6	76.405	75.600
21)	B Endrin ke...	7.080	7.891	136.2E6	736.0E6	76.575	75.807
22)	Mirex	7.261	8.355	95937688	493.4E6	76.285	75.528

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061322\
Data File : PD070264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2022 17:26
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 04:27:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061322.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 04:26:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

