

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD061324\  
 Data File : PD083684.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 13 Jun 2024 11:47  
 Operator : AR\AJ  
 Sample : P2648-07  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 BHBW1

Manual Integrations  
 APPROVED

Reviewed By : Abdul Mirza 06/14/2024  
 Supervised By : Ankita Jodhani 06/14/2024

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jun 14 02:20:31 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD060724CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Jun 07 04:03:14 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm


| Compound                    | RT#1  | RT#2  | Resp#1   | Resp#2   | ng/ml    | ng/ml      |
|-----------------------------|-------|-------|----------|----------|----------|------------|
| -----                       |       |       |          |          |          |            |
| System Monitoring Compounds |       |       |          |          |          |            |
| 1) SA Tetrachlo...          | 3.560 | 2.769 | 9164393  | 43089137 | 7.021m   | 7.905m     |
| 27) SA Decachlor...         | 9.112 | 7.897 | 26439496 | 77188007 | 13.597m  | 12.712     |
| Target Compounds            |       |       |          |          |          |            |
| 10) B trans-Chl...          | 5.973 | 4.965 | 21511180 | 61565058 | 10.565m  | 8.476m     |
| 11) B cis-Chlor...          | 6.055 | 5.029 | 32168905 | 53529921 | 15.448   | 7.271m#    |
| 12) B 4,4'-DDE              | 6.226 | 5.219 | 223.3E6  | 954.6E6  | 107.649  | 134.323    |
| 13) MA Dieldrin             | 6.379 | 5.340 | 20562498 | 127.2E6  | 9.945m   | 17.460m#   |
| 15) B Endosulfa...          | 6.836 | 5.923 | 15218615 | 19296510 | 8.461m   | 3.041m#    |
| 16) A 4,4'-DDD              | 6.741 | 5.773 | 18062012 | 53085734 | 12.326m  | 9.417m     |
| 17) MA 4,4'-DDT             | 7.061 | 6.023 | 2222.0E6 | 5914.5E6 | 1475.026 | 1044.870 # |
| -----                       |       |       |          |          |          |            |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD061324\  
Data File : PD083684.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 13 Jun 2024 11:47  
Operator : AR\AJ  
Sample : P2648-07  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jun 14 02:20:31 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD060724CLP.M  
Quant Title : GC Extractables  
QLast Update : Fri Jun 07 04:03:14 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Mirza 06/14/2024  
Supervised By: Ankita Jodhani 06/14/2024

Manual Integrations  
APPROVED

Instrument :  
ECD\_D  
Client Sampled :  
BHBW1