

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061521\
 Data File : PD063875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2021 21:58
 Operator : AR\AJ
 Sample : M2596-17
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 02:27:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061421CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 15 03:10:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.438	3.971	18141095	166.5E6	11.670	10.513
27)	SA Decachlor...	8.192	9.026	43143370	318.9E6	24.924	22.995
Target Compounds							
3)	MA gamma-BHC...	4.161	4.622	17929538	72782579	7.878	3.190m#
5)	MB Aldrin	4.703	5.453	2821713	9998090	1.392m	0.454 #
7)	B delta-BHC	4.609	5.023	2667043	20326246	1.236	0.868m#
10)	B trans-Chl...	5.374	6.071	16330598	217.7E6	8.398m	10.348m
11)	B cis-Chlor...	5.432	6.148	11545024	213.4E6	6.076m	10.598 #
12)	B 4,4'-DDE	5.596	6.301	25536028	205.0E6	12.934	10.146
13)	MA Dieldrin	5.729	6.456	27662016	57592212	13.218m	2.753m#
16)	A 4,4'-DDD	6.114	6.794	14531108	341.6E6	8.801	20.291 #
17)	MA 4,4'-DDT	6.354	7.093	134.0E6	1196.3E6	79.617	72.667

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061521\
Data File : PD063875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2021 21:58
Operator : AR\AJ
Sample : M2596-17
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 02:27:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061421CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 15 03:10:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

