

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061723\
 Data File : PD075833.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jun 2023 10:22
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/19/2023
 Supervised By :Ankita Jodhani 06/19/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 11:02:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 16 10:43:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.767	114.2E6	56438462	53.634	52.020
28)	SA Decachlor...	9.090	7.913	133.3E6	51163099	52.625	52.325
Target Compounds							
2)	A alpha-BHC	4.004	3.270	203.9E6	84662426	55.215	52.922
3)	MA gamma-BHC...	4.337	3.599	186.8E6	78440076	55.525	53.236
4)	MA Heptachlor	4.929	3.940	195.6E6	78847627	54.544	51.913
5)	MB Aldrin	5.271	4.220	191.7E6	77711447	54.713	52.114
6)	B beta-BHC	4.533	3.896	76647774	32955339	53.312	51.702
7)	B delta-BHC	4.782	4.127	186.9E6	77630925	56.239	53.576
8)	B Heptachlo...	5.699	4.723	172.9E6	70869935	53.651	51.593
9)	A Endosulfan I	6.086	5.093	160.7E6	66413808	53.585	51.959
10)	B gamma-Chl...	5.956	4.973	175.7E6	71494246	54.542	52.105
11)	B alpha-Chl...	6.035	5.038	173.6E6	71070260	53.695	51.809
12)	B 4,4'-DDE	6.207	5.227	158.2E6	60651612	55.337	54.339
13)	MA Dieldrin	6.362	5.358	169.9E6	70574695	54.372	52.310
14)	MA Endrin	6.592	5.634	142.5E6	62377534	53.291	51.256
15)	B Endosulfa...	6.812	5.928	136.3E6	56703114	52.753	52.559
16)	A 4,4'-DDD	6.726	5.782	126.5E6	49408356	55.594	56.065
17)	MA 4,4'-DDT	7.041	6.034	128.2E6	47412978	56.509	58.369
18)	B Endrin al...	6.941	6.108	106.5E6	44976507	53.056m	51.966
19)	B Endosulfa...	7.177	6.331	134.7E6	57070954	53.486	52.814
20)	A Methoxychlor	7.517	6.609	69147715	25597785	53.421	55.306
21)	B Endrin ke...	7.664	6.837	151.8E6	65675905	54.474	54.056
22)	Mirex	8.139	7.019	113.7E6	53772011	53.070	51.298

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061723\
Data File : PD075833.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jun 2023 10:22
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/19/2023

Supervised By :Ankita Jodhani 06/19/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 11:02:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
Quant Title : GC Extractables
QLast Update : Fri Jun 16 10:43:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

