

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061723\
 Data File : PD075868.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jun 2023 19:03
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/19/2023
 Supervised By :Ankita Jodhani 06/19/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 22:12:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 16 10:43:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.768	109.2E6	54628823	51.243	50.352
28)	SA Decachlor...	9.088	7.914	116.0E6	46585675	45.796	47.644
Target Compounds							
2)	A alpha-BHC	4.003	3.270	195.2E6	82120878	52.872	51.333
3)	MA gamma-BHC...	4.336	3.600	177.8E6	75405338	52.849	51.177
4)	MA Heptachlor	4.927	3.940	179.3E6	74501333	49.997	49.051
5)	MB Aldrin	5.270	4.220	177.0E6	74306490	50.497	49.830
6)	B beta-BHC	4.532	3.896	73595650	32923056	51.189	51.651
7)	B delta-BHC	4.780	4.127	178.7E6	76788120	53.758	52.995
8)	B Heptachlo...	5.698	4.722	157.8E6	67969405	48.978	49.481
9)	A Endosulfan I	6.085	5.093	145.7E6	62621134	48.571	48.991
10)	B gamma-Chl...	5.954	4.973	159.2E6	68794491	49.420	50.138
11)	B alpha-Chl...	6.034	5.037	157.3E6	67776032	48.647	49.408
12)	B 4,4'-DDE	6.206	5.226	144.1E6	61420433	50.408	55.028
13)	MA Dieldrin	6.360	5.358	151.4E6	67072071	48.450	49.714
14)	MA Endrin	6.591	5.634	126.8E6	58572723	47.407	48.129
15)	B Endosulfa...	6.810	5.928	121.3E6	53714601	46.967	49.789
16)	A 4,4'-DDD	6.724	5.781	117.3E6	52004572	51.536	59.011
17)	MA 4,4'-DDT	7.040	6.033	111.9E6	46758962	49.329	57.564
18)	B Endrin al...	6.939	6.108	95897306	43007234	47.791m	49.691
19)	B Endosulfa...	7.176	6.331	119.2E6	54606248	47.301	50.533
20)	A Methoxychlor	7.515	6.609	60844424	26242795	47.006	56.699
21)	B Endrin ke...	7.662	6.837	134.1E6	61947139	48.143	50.987
22)	Mirex	8.137	7.020	98510582	48340972	45.978	46.117

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061723\
Data File : PD075868.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jun 2023 19:03
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/19/2023
Supervised By :Ankita Jodhani 06/19/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 22:12:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
Quant Title : GC Extractables
QLast Update : Fri Jun 16 10:43:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

