

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061924\
 Data File : PD083756.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2024 17:51
 Operator : AR\AJ
 Sample : P2899-15MS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0AB6MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/20/2024
 Supervised By :Ankita Jodhani 06/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 22:40:15 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 07 04:03:14 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.561	2.771	10435120	42339243	7.994	7.767m
27)	SA Decachlor...	9.110	7.894	14153940	42064098	7.279m	6.927m
Target Compounds							
2)	A alpha-BHC	4.017	3.271	83658731	307.6E6	37.695m	36.087
3)	MA gamma-BHC...	4.351	3.598	82232721	279.8E6	37.798m	35.832m
4)	MA Heptachlor	4.942	3.937	84300856	303.4E6	34.701m	38.758
5)	MB Aldrin	5.287	4.216	80123534	269.4E6	33.726	34.863
6)	B beta-BHC	4.548	3.896	33460667	145.1E6	33.202m	40.655
7)	B delta-BHC	4.798	4.125	91009880	289.1E6	34.645	33.901
8)	B Heptachlo...	5.715	4.718	69236390	252.0E6	33.522	34.443
9)	A Endosulfan I	6.102	5.087	72400938	235.9E6	37.023	35.685
10)	B trans-Chl...	5.971	4.967	74596702	265.1E6	36.636m	36.495
11)	B cis-Chlor...	6.051	5.031	76580781	263.7E6	36.775	35.821
12)	B 4,4'-DDE	6.223	5.218	154.0E6	470.9E6	74.228	66.259
13)	MA Dieldrin	6.378	5.350	162.3E6	525.8E6	78.502	72.179
14)	MA Endrin	6.609	5.625	141.4E6	479.3E6	83.680	74.112
15)	B Endosulfa...	6.827	5.918	122.2E6	387.2E6	67.962m	61.014m
16)	A 4,4'-DDD	6.738	5.772	179.6E6	316.7E6	122.601m	56.181 #
17)	MA 4,4'-DDT	7.057	6.022	132.8E6	383.6E6	88.127	67.770
18)	B Endrin al...	6.959	6.098	106.9E6	281.7E6	80.135	58.803 #
19)	B Endosulfa...	7.194	6.320	133.4E6	386.7E6	77.951	62.585
20)	A Methoxychlor	7.532	6.596	336.9E6	889.2E6	389.441m	305.962m
21)	B Endrin ke...	7.680	6.824	146.0E6	429.5E6	78.343	63.883m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061924\
Data File : PD083756.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2024 17:51
Operator : AR\AJ
Sample : P2899-15MS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 22:40:15 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
COAB6MS

Manual Integrations
APPROVED

Reviewed By: Abdul Mirza 06/20/2024
Supervised By: Ankita Jodhani 06/20/2024