

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
 Data File : PD070383.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2022 09:33
 Operator : AR\AJ
 Sample : PB145661BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS661

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 03:04:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 14 16:58:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.400	4.120	30727925	277.9E6	21.187	20.572
2)	SA Decachlor...	8.127	9.277	61594821	278.2E6	38.456	38.554
Target Compounds							
2)	A alpha-BHC	3.835	4.513	10150885	95717602	4.803	4.890
3)	MA gamma-BHC...	4.116	4.802	10339604	95977118	5.149	5.263
4)	MA Heptachlor	4.410	5.325	9963213	90781905	5.062	5.067
5)	MB Aldrin	4.654	5.635	10496071	91986038	5.503	5.534
6)	B beta-BHC	4.366	4.966	4973418	43882528	5.556	5.547
7)	B delta-BHC	4.567	5.189	7890977	52523766	3.910	4.071
8)	B Heptachlo...	5.095	6.022	10093715	73195614	5.762	6.273
9)	A Endosulfan I	5.431	6.383	9321732	74644795	5.542	5.535
10)	B trans-Chl...	5.320	6.258	10221193	79819931	5.683	5.523
11)	B cis-Chlor...	5.378	6.332	10020810	79439605	5.666	5.658
12)	B 4,4'-DDE	5.544	6.484	18329860	144.9E6	10.858	11.146
13)	MA Dieldrin	5.674	6.641	19832887	145.9E6	10.455	10.651
14)	MA Endrin	5.930	6.860	16579118	115.7E6	9.742	10.054
15)	B Endosulfa...	6.204	7.066	17187131	121.0E6	10.591	10.999
16)	A 4,4'-DDD	6.059	6.974	16092828	123.7E6	10.629	11.200
17)	MA 4,4'-DDT	6.297	7.276	12470804	90451755	8.570	8.797
18)	B Endrin al...	6.373	7.189	15478661	102.6E6	12.443	12.200
19)	B Endosulfa...	6.585	7.414	15459123	96796203	9.760	9.793
20)	A Methoxychlor	6.847	7.730	40063604	221.1E6	47.335	46.832
21)	B Endrin ke...	7.075	7.888	19877019	114.6E6	11.146	11.249

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070383.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 09:33
Operator : AR\AJ
Sample : PB145661BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:04:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

