

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 11:42
Operator : AR\AJ
Sample : N3374-09
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW4T9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/21/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(1) Tetrachloro-m-xylene (SA)

3.399min 14.706 ng/ml

response 21327920

(1) Tetrachloro-m-xylene #2 (SA)

4.120min 8.068 ng/ml

response 108995862

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 11:42
Operator : AR\AJ
Sample : N3374-09
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW4T9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/21/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(1) Tetrachloro-m-xylene (SA)

3.399min 14.706 ng/ml

response 21327920

(1) Tetrachloro-m-xylene #2 (SA)

4.119min 11.260 ng/ml m

response 152103353

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 11:42
Operator : AR\AJ
Sample : N3374-09
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW4T9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/21/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(13) Dieldrin (MA)

5.677min 1.452 ng/ml
response 2754182

(13) Dieldrin #2 (MA)

6.619min 1.413 ng/ml
response 19356872

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 11:42
Operator : AR\AJ
Sample : N3374-09
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW4T9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/21/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(13) Dieldrin (MA)

5.675min 0.543 ng/ml m
response 1029215

(13) Dieldrin #2 (MA)

6.618min 3.384 ng/ml m
response 46341433

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 11:42
Operator : AR\AJ
Sample : N3374-09
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW4T9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/21/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(14) Endrin (MA)

5.952min 3.742 ng/ml

response 6367531

(14) Endrin #2 (MA)

6.883min 2.962 ng/ml

response 34075217

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 11:42
Operator : AR\AJ
Sample : N3374-09
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW4T9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/21/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(14) Endrin (MA)

5.951min 4.592 ng/ml m

response 7814191

(14) Endrin #2 (MA)

6.883min 2.962 ng/ml

response 34075217

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 11:42
Operator : AR\AJ
Sample : N3374-09
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW4T9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/21/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(16) 4,4'-DDD (A)

6.073min 1.806 ng/ml
response 2734589

(16) 4,4'-DDD #2 (A)

6.972min 0.756 ng/ml
response 8344644

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062022\
Data File : PD070391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 11:42
Operator : AR\AJ
Sample : N3374-09
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
EW4T9

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 06/21/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061422CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 14 16:58:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(16) 4,4'-DDD (A)

6.071min 1.516 ng/ml m
response 2295569

(16) 4,4'-DDD #2 (A)

6.971min 0.670 ng/ml m
response 7402322