

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062221\
 Data File : PD063942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2021 12:29
 Operator : AR\AJ
 Sample : INDB312
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB3012

Manual Integrations
 APPROVED

Ankita
 6/23/2021 11:10:39 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 02:37:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061421CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 21 04:24:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.443	3.971	30681088	306.8E6	19.737	19.372
27)	SA Decachlor...	8.200	9.028	69962340	587.6E6	40.418	42.374
Target Compounds							
5)	MB Aldrin	4.710	5.455	43280761	459.8E6	21.354m	20.869
6)	B beta-BHC	4.415	4.815	22610882	220.6E6	23.286	22.622
7)	B delta-BHC	4.618	5.031	46026646	483.4E6	21.323	20.633
8)	B Heptachlo...	5.154	5.839	40239593	468.2E6	21.220	22.756m
10)	B trans-Chl...	5.382	6.076	41679881	443.3E6	21.433	21.072
11)	B cis-Chlor...	5.440	6.149	39801231	398.3E6	20.949	19.777
12)	B 4,4'-DDE	5.602	6.305	85329584	789.3E6	43.219m	39.064
15)	B Endosulfa...	6.269	6.882	68923965	650.8E6	42.178	40.221
18)	B Endrin al...	6.437	7.006	59140396	576.1E6	41.424	40.951
19)	B Endosulfa...	6.651	7.229	65438994	650.6E6	42.154	39.531
21)	B Endrin ke...	7.143	7.700	78854141	644.4E6	41.566	40.606

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062221\
Data File : PD063942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2021 12:29
Operator : AR\AJ
Sample : INDB312
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 02:37:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061421CLP.M
Quant Title : GC Extractables
QLast Update : Mon Jun 21 04:24:00 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

6/23/2021 11:10:39 AM

Instrument :
ECD_D
ClientSampled :
INDB3012