

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062320\
Data File : PD058379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2020 10:30
Operator : AJ\MA
Sample : L3077-12
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK-SPIKE

Manual Integrations
APPROVED

mohammad
6/24/2020 11:45:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 00:46:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062220CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 23 04:50:18 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(5) Aldrin (MB)

4.418min 20.856 ng/ml

response 41157769

(5) Aldrin #2 (MB)

5.330min 19.905 ng/ml

response 56071979

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062320\
Data File : PD058379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2020 10:30
Operator : AJ\MA
Sample : L3077-12
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK-SPIKE

Manual Integrations
APPROVED

mohammad
6/24/2020 11:45:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 00:46:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062220CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 23 04:50:18 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(5) Aldrin (MB)

4.417min 19.514 ng/ml m
response 38509180

(5) Aldrin #2 (MB)

5.330min 19.905 ng/ml
response 56071979

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062320\
Data File : PD058379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2020 10:30
Operator : AJ\MA
Sample : L3077-12
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK-SPIKE

Manual Integrations
APPROVED

mohammad
6/24/2020 11:45:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 00:46:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062220CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 23 04:50:18 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(13) Dieldrin (MA)

5.416min 40.549 ng/ml

response 60556092

(13) Dieldrin #2 (MA)

6.321min 40.058 ng/ml

response 106286334

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062320\
 Data File : PD058379.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jun 2020 10:30
 Operator : AJ\MA
 Sample : L3077-12
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC2-BLANK-SPIKE

Manual Integrations
 APPROVED

mohammad
 6/24/2020 11:45:50 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 00:46:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062220CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 23 04:50:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(13) Dieldrin (MA)
 5.414min 36.766 ng/ml m
 response 54906529

(13) Dieldrin #2 (MA)
 6.321min 40.058 ng/ml
 response 106286334

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062320\
Data File : PD058379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2020 10:30
Operator : AJ\MA
Sample : L3077-12
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK-SPIKE

Manual Integrations
APPROVED

mohammad
6/24/2020 11:45:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 00:46:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062220CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 23 04:50:18 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(14) Endrin (MA)

5.666min 41.259 ng/ml

response 52990673

(14) Endrin #2 (MA)

6.536min 41.962 ng/ml

response 91117049

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062320\
Data File : PD058379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2020 10:30
Operator : AJ\MA
Sample : L3077-12
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK-SPIKE

Manual Integrations
APPROVED

mohammad
6/24/2020 11:45:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 00:46:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062220CLP.M
Quant Title : GC Extractables
QLast Update : Tue Jun 23 04:50:18 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(14) Endrin (MA)
5.665min 37.758 ng/ml m
response 48493534

(14) Endrin #2 (MA)
6.536min 41.962 ng/ml
response 91117049