

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062321\
 Data File : PD063948.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jun 2021 09:47
 Operator : AR\AJ
 Sample : M2682-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BGDF3

Manual Integrations
 APPROVED

Ankita
 6/25/2021 8:51:17 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 04:56:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061421CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 21 04:24:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.437	3.968	11869806	106.3E6	7.636m	6.714m
27)	SA Decachlor...	8.187	9.021	60295604	195.0E6	34.834	14.062m#
Target Compounds							
3)	MA gamma-BHC...	4.160	4.623	15754420	160.5E6	6.923	7.034m
7)	B delta-BHC	4.605	5.026	6295822	22029484	2.917m	0.940 #
12)	B 4,4'-DDE	5.614	6.293	23048930	67251910	11.674m	3.329m#
14)	MA Endrin	5.981	6.688	7311501	180.3E6	4.043	10.305m#
15)	B Endosulfa...	6.262	6.870	15266503	53827073	9.342	3.327m#
16)	A 4,4'-DDD	6.114	6.786	5320249	125.0E6	3.222m	7.428m#
19)	B Endosulfa...	6.658	7.225	22599047	145.7E6	14.558m	8.854m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062321\
Data File : PD063948.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2021 09:47
Operator : AR\AJ
Sample : M2682-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 04:56:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061421CLP.M
Quant Title : GC Extractables
QLast Update : Mon Jun 21 04:24:00 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

6/25/2021 8:51:17 AM

Ankita

Instrument :
ECD_D
Client Sampled :
BGDF3