

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062322\
 Data File : PD070474.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2022 21:19
 Operator : AR\AJ
 Sample : INDB201
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB2097

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/23/2022
 Supervised By : Ankita Jodhani 06/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 10:45:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062322CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 23 10:45:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.400	4.121	15542547	152.4E6	10.207	10.343
27)	SA Decachlor...	8.127	9.277	36011743	174.8E6	20.609	20.355
Target Compounds							
5)	MB Aldrin	4.654	5.636	21636434	190.1E6	10.193	10.196
6)	B beta-BHC	4.365	4.967	10706404	95973305	10.234	10.343
7)	B delta-BHC	4.566	5.189	22868365	95138826	10.256	11.001m
8)	B Heptachlo...	5.095	6.023	20576990	81788220	10.243	11.735
10)	B trans-Chl...	5.319	6.260	20927923	167.5E6	10.260	10.256
11)	B cis-Chlor...	5.378	6.334	20825132	164.2E6	10.202	10.229
12)	B 4,4'-DDE	5.544	6.485	39449623	303.3E6	20.622	20.496
15)	B Endosulfa...	6.203	7.068	34947324	243.4E6	20.611	20.515
18)	B Endrin al...	6.372	7.191	31957601	215.6E6	20.654	20.496
19)	B Endosulfa...	6.585	7.416	37337373	242.5E6	20.776	20.539
21)	B Endrin ke...	7.075	7.889	41322641	247.8E6	20.618	20.413

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062322\
Data File : PD070474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2022 21:19
Operator : AR\AJ
Sample : INDB201
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 10:45:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062322CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 23 10:45:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Mirza 06/23/2022
Supervised By: Ankita Jodhani 06/23/2022

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
INDB2097