

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062322\
 Data File : PD070475.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2022 21:33
 Operator : AR\AJ
 Sample : INDA301
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3098

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 11:44:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062322CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 23 11:43:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.400	4.121	30590510	291.5E6	20.000	20.000
27)	SA Decachlor...	8.126	9.277	65153749	303.0E6	40.000	40.000
Target Compounds							
2)	A alpha-BHC	3.835	4.514	44799602	416.5E6	20.000	20.000
3)	MA gamma-BHC...	4.115	4.802	42840339	378.4E6	20.000	20.000
4)	MA Heptachlor	4.409	5.326	42715747	358.2E6	20.000	20.000
9)	A Endosulfan I	5.431	6.384	36205202	275.4E6	20.000	20.000
13)	MA Dieldrin	5.673	6.642	82461288	564.7E6	40.000	40.000
14)	MA Endrin	5.930	6.861	73432569	453.6E6	40.000	40.000
16)	A 4,4'-DDD	6.058	6.976	64452177	447.9E6	40.000	40.000
17)	MA 4,4'-DDT	6.297	7.278	64392961	422.0E6	40.000	40.000
20)	A Methoxychlor	6.846	7.732	185.4E6	968.3E6	200.000	200.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062322\
Data File : PD070475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2022 21:33
Operator : AR\AJ
Sample : INDA301
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 11:44:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062322CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 23 11:43:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

