

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062423\
 Data File : PD076056.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2023 18:01
 Operator : AR\AJ
 Sample : 03336-16
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 WC-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 22:23:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 16 10:43:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.767	25462639	13141028	11.954	12.112
28)	SA Decachlor...	9.083	7.911	33012154	14845470	13.031	15.183
Target Compounds							
4)	MA Heptachlor	4.943f	0.000	305359	0	0.085	N.D. #
7)	B delta-BHC	4.771	0.000	1750706	0	0.527	N.D. #
10)	B gamma-Chl...	5.948	0.000	106944	0	0.033	N.D. #
15)	B Endosulfa...	6.836f	0.000	146612	0	0.057	N.D. #
20)	A Methoxychlor	7.549f	0.000	50114	0	0.039	N.D. #
21)	B Endrin ke...	7.689f	0.000	637715	0	0.229	N.D. #
25)	Chlordane-3	5.948	0.000	106944	0	0.206	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062423\
Data File : PD076056.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2023 18:01
Operator : AR\AJ
Sample : 03336-16
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
WC-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 22:23:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
Quant Title : GC Extractables
QLast Update : Fri Jun 16 10:43:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 25462639
Conc: 11.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
WC-1




#1 Tetrachloro-m-xylene

R.T.: 2.767 min
Delta R.T.: 0.000 min
Response: 13141028
Conc: 12.11 ng/ml




#4 Heptachlor

R.T.: 4.943 min
Delta R.T.: 0.016 min
Response: 305359
Conc: 0.09 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T. : 3.940 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 4.771 min
Delta R.T.: -0.010 min
Response: 1750706
Conc: 0.53 ng/ml

Instrument :
ECD_D
ClientSampleId :
WC-1




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.127 min
Response: 0
Conc: N.D.




#10 gamma-Chlordane

R.T.: 5.948 min
Delta R.T.: -0.007 min
Response: 106944
Conc: 0.03 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T. : 4.973 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.836 min
Delta R.T.: 0.026 min
Response: 146612
Conc: 0.06 ng/ml

Instrument :
ECD_D
ClientSampleId :
WC-1




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 5.928 min
Response: 0
Conc: N.D.




#20 Methoxychlor

R.T.: 7.549 min
Delta R.T.: 0.034 min
Response: 50114
Conc: 0.04 ng/ml



#20 Methoxychlor


R.T.: 0.000 min
Exp R.T. : 6.608 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 7.689 min
Delta R.T.: 0.027 min
Response: 637715
Conc: 0.23 ng/ml

Instrument :
ECD_D
ClientSampleId :
WC-1




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T. : 6.837 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.948 min
Delta R.T.: -0.008 min
Response: 106944
Conc: 0.21 ng/ml



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T. : 4.974 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 9.083 min
Delta R.T.: -0.006 min
Response: 33012154
Conc: 13.03 ng/ml

Instrument :
ECD_D
ClientSampleId :
WC-1



#28 Decachlorobiphenyl

R.T.: 7.911 min
Delta R.T.: -0.002 min
Response: 14845470
Conc: 15.18 ng/ml