

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062425\
 Data File : PD089106.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2025 14:04
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 01:53:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	142.5E6	832.9E6	49.951	49.132
28)	SA Decachlor...	9.072	8.072	185.6E6	919.6E6	47.250	46.516
Target Compounds							
2)	A alpha-BHC	3.999	3.393	301.1E6	1318.4E6	51.658	49.198
3)	MA gamma-BHC...	4.330	3.729	285.8E6	1210.0E6	51.203	48.876
4)	MA Heptachlor	4.929	4.082	270.5E6	1184.9E6	49.610	47.322
5)	MB Aldrin	5.271	4.369	270.1E6	1180.7E6	50.836	48.756
6)	B beta-BHC	4.515	4.025	109.1E6	520.0E6	49.420	48.393
7)	B delta-BHC	4.764	4.262	274.4E6	1211.5E6	53.468	48.705
8)	B Heptachlo...	5.690	4.872	237.8E6	1057.1E6	49.289	48.272
9)	A Endosulfan I	6.074	5.247	223.6E6	1007.3E6	49.607	48.330
10)	B gamma-Chl...	5.945	5.126	242.2E6	1146.0E6	50.588	47.906
11)	B alpha-Chl...	6.026	5.190	240.8E6	1095.2E6	49.775	47.830
12)	B 4,4'-DDE	6.196	5.375	217.6E6	1100.7E6	49.898	48.034
13)	MA Dieldrin	6.347	5.513	240.4E6	1118.7E6	50.113	48.204
14)	MA Endrin	6.574	5.789	195.7E6	990.9E6	47.430	45.634
15)	B Endosulfa...	6.786	6.080	202.0E6	967.3E6	50.031	47.894
16)	A 4,4'-DDD	6.705	5.929	171.9E6	924.4E6	50.072	48.322
17)	MA 4,4'-DDT	7.021	6.183	175.6E6	921.9E6	46.366	45.458
18)	B Endrin al...	6.915	6.258	146.6E6	714.3E6	47.855	46.649
19)	B Endosulfa...	7.149	6.482	187.2E6	929.0E6	48.783	47.345
20)	A Methoxychlor	7.493	6.754	88237776	470.7E6	43.668	43.708
21)	B Endrin ke...	7.629	6.991	199.5E6	1035.0E6	48.988	47.914
22)	Mirex	8.113	7.185	143.4E6	775.5E6	47.168	46.496

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062425\
Data File : PD089106.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2025 14:04
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 01:53:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

