

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062523\
 Data File : PD076136.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2023 01:04
 Operator : AR\AJ
 Sample : 03376-01
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 FB-GW-062123

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 05:35:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 16 10:43:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.769	33002914	15809208	15.493	14.572
28)	SA Decachlor...	9.086	7.910	37064091	14992681	14.630	15.333
Target Compounds							
3)	MA gamma-BHC...	4.363f	0.000	51219	0	0.015	N.D. #
5)	MB Aldrin	5.282	0.000	269123	0	0.077	N.D. #
6)	B beta-BHC	4.523	0.000	563018	0	0.392	N.D. #
7)	B delta-BHC	4.772	0.000	3461683	0	1.042	N.D. #
8)	B Heptachlo...	5.691	0.000	2326640	0	0.722	N.D. #
9)	A Endosulfan I	6.083	0.000	517532	0	0.173	N.D. #
10)	B gamma-Chl...	5.978f	0.000	679171	0	0.211	N.D. #
13)	MA Dieldrin	6.367	0.000	283882	0	0.091	N.D. #
15)	B Endosulfa...	6.835f	0.000	294748	0	0.114	N.D. #
21)	B Endrin ke...	7.690f	0.000	1310839	0	0.471	N.D. #
25)	Chlordane-3	5.978f	0.000	679171	0	1.308	N.D. #
27)	Chlordane-5	6.885	0.000	128302	0	1.287	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062523\
Data File : PD076136.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2023 01:04
Operator : AR\AJ
Sample : 03376-01
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
FB-GW-062123

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 05:35:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
Quant Title : GC Extractables
QLast Update : Fri Jun 16 10:43:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.547 min
Delta R.T.: 0.000 min
Response: 33002914
Conc: 15.49 ng/ml

Instrument :
ECD_D
ClientSampleId :
FB-GW-062123




#1 Tetrachloro-m-xylene

R.T.: 2.769 min
Delta R.T.: 0.002 min
Response: 15809208
Conc: 14.57 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.363 min
Delta R.T.: 0.027 min
Response: 51219
Conc: 0.02 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 0.000 min
Exp R.T. : 3.599 min
Response: 0
Conc: N.D.



#5 Aldrin


R.T.: 5.282 min
Delta R.T.: 0.012 min
Response: 269123
Conc: 0.08 ng/ml

Instrument :
ECD_D
ClientSampleId :
FB-GW-062123




#5 Aldrin

R.T.: 0.000 min
Exp R.T. : 4.220 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 4.523 min
Delta R.T.: -0.009 min
Response: 563018
Conc: 0.39 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 3.896 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 4.772 min
Delta R.T.: -0.009 min
Response: 3461683
Conc: 1.04 ng/ml

Instrument :
ECD_D
ClientSampleId :
FB-GW-062123




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.127 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.691 min
Delta R.T.: -0.006 min
Response: 2326640
Conc: 0.72 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T. : 4.723 min
Response: 0
Conc: N.D.



#9 Endosulfan I


R.T.: 6.083 min
Delta R.T.: -0.002 min
Response: 517532
Conc: 0.17 ng/ml

Instrument :
ECD_D
ClientSampleId :
FB-GW-062123




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T. : 5.093 min
Response: 0
Conc: N.D.




#10 gamma-Chlordane

R.T.: 5.978 min
Delta R.T.: 0.023 min
Response: 679171
Conc: 0.21 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T. : 4.973 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 6.367 min
Delta R.T.: 0.007 min
Response: 283882
Conc: 0.09 ng/ml

Instrument :
ECD_D
ClientSampleId :
FB-GW-062123




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 5.358 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 6.835 min
Delta R.T.: 0.024 min
Response: 294748
Conc: 0.11 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T. : 5.928 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 7.690 min
Delta R.T.: 0.027 min
Response: 1310839
Conc: 0.47 ng/ml

Instrument :
ECD_D
ClientSampleId :
FB-GW-062123




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T. : 6.837 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.978 min
Delta R.T.: 0.022 min
Response: 679171
Conc: 1.31 ng/ml



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T. : 4.974 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 6.885 min
Delta R.T.: -0.006 min
Response: 128302
Conc: 1.29 ng/ml

Instrument :
ECD_D
ClientSampleId :
FB-GW-062123




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 5.934 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 9.086 min
Delta R.T.: -0.003 min
Response: 37064091
Conc: 14.63 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.910 min
Delta R.T.: -0.002 min
Response: 14992681
Conc: 15.33 ng/ml