

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062523\
 Data File : PD076144.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2023 03:20
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/26/2023
 Supervised By : Ankita Jodhani 06/26/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 05:38:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 16 10:43:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.767	101.2E6	50600757	47.491	46.640
28)	SA Decachlor...	9.084	7.911	105.5E6	45387331	41.661	46.418
Target Compounds							
2)	A alpha-BHC	4.002	3.269	177.7E6	76477186	48.118	47.806
3)	MA gamma-BHC...	4.336	3.599	163.0E6	71068615	48.450	48.233
4)	MA Heptachlor	4.926	3.939	165.2E6	70674417	46.046	46.531
5)	MB Aldrin	5.270	4.219	165.0E6	73268005	47.100	49.134
6)	B beta-BHC	4.532	3.896	67845188	30987456	47.189	48.615
7)	B delta-BHC	4.780	4.126	163.9E6	73480716	49.310	50.712
8)	B Heptachlo...	5.696	4.721	150.4E6	64972168	46.669	47.299
9)	A Endosulfan I	6.084	5.091	138.3E6	58522598	46.114	45.785
10)	B gamma-Chl...	5.953	4.972	152.5E6	66553199	47.325	48.504
11)	B alpha-Chl...	6.033	5.036	149.7E6	65733355	46.313	47.919
12)	B 4,4'-DDE	6.205	5.225	135.9E6	60310057	47.533	54.033
13)	MA Dieldrin	6.359	5.357	144.1E6	66060309	46.103	48.964
14)	MA Endrin	6.590	5.632	113.8E6	55797004	42.553	45.848
15)	B Endosulfa...	6.808	5.926	113.9E6	53115396	44.095	49.233
16)	A 4,4'-DDD	6.723	5.780	112.7E6	52666685	49.528	59.762
17)	MA 4,4'-DDT	7.038	6.031	104.1E6	46113147	45.863	56.769
18)	B Endrin al...	6.937	6.106	90005869	42205133	44.855m	48.764
19)	B Endosulfa...	7.174	6.328	113.5E6	54314478	45.038	50.263
20)	A Methoxychlor	7.513	6.606	55596293	26041678	42.952	56.265 #
21)	B Endrin ke...	7.660	6.835	130.5E6	62515406	46.829	51.455
22)	Mirex	8.135	7.017	92312467	48564154	43.085	46.330

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062523\
Data File : PD076144.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2023 03:20
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/26/2023

Supervised By :Ankita Jodhani 06/26/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 05:38:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
Quant Title : GC Extractables
QLast Update : Fri Jun 16 10:43:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

