

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062624\
 Data File : PD083823.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2024 16:21
 Operator : AR\AJ
 Sample : INDA301
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3012

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 16:44:04 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062624CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 26 16:43:46 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.562	2.772	20550129	109.2E6	20.000	20.000
27)	SA Decachlor...	9.117	7.897	85037968	269.6E6	40.000	40.000
Target Compounds							
2)	A alpha-BHC	4.019	3.271	37151723	183.0E6	20.000	20.000
3)	MA gamma-BHC...	4.353	3.599	36756393	173.9E6	20.000	20.000
4)	MA Heptachlor	4.945	3.937	40032927	170.1E6	20.000	20.000
9)	A Endosulfan I	6.104	5.086	37241653	138.8E6	20.000	20.000
13)	MA Dieldrin	6.380	5.351	80401966	316.2E6	40.000	40.000
14)	MA Endrin	6.612	5.626	66280953	270.5E6	40.000	40.000
16)	A 4,4'-DDD	6.744	5.773	58031657	243.3E6	40.000	40.000
17)	MA 4,4'-DDT	7.060	6.023	60655307	240.4E6	40.000	40.000
20)	A Methoxychlor	7.537	6.598	178.8E6	612.1E6	200.000	200.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062624\
Data File : PD083823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2024 16:21
Operator : AR\AJ
Sample : INDA301
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 16:44:04 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062624CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jun 26 16:43:46 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

