

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062724\
 Data File : PD083840.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2024 15:36
 Operator : AR\AJ
 Sample : INDA101
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA1024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 18:20:47 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062724CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 27 18:05:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.561	2.772	4743592	27947270	4.793	5.000
27)	SA Decachlor...	9.116	7.897	21236059	71166701	10.214	10.734
Target Compounds							
2)	A alpha-BHC	4.018	3.271	8330863	42629017	4.513	4.732
3)	MA gamma-BHC...	4.353	3.599	8342442	40027259	4.590	4.734
4)	MA Heptachlor	4.945	3.937	9187343	40770326	4.742	4.845
9)	A Endosulfan I	6.103	5.086	8715336	34682729	4.815	4.973
13)	MA Dieldrin	6.380	5.350	18271774	75677966	9.296	9.779
14)	MA Endrin	6.610	5.625	13954775	60493290	9.153	9.646
16)	A 4,4'-DDD	6.743	5.772	13179314	59324167	9.201	9.822
17)	MA 4,4'-DDT	7.059	6.023	12490560	50149952	8.719	8.853
20)	A Methoxychlor	7.537	6.598	40438839	149.0E6	48.143	52.548

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062724\
Data File : PD083840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2024 15:36
Operator : AR\AJ
Sample : INDA101
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 18:20:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062724CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 27 18:05:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

