

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062725\
 Data File : PD089188.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2025 12:53
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 05:37:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.553	2.881	164.9E6	968.6E6	57.795	57.135
28)	SA Decachlor...	9.074	8.072	186.4E6	882.4E6	47.460	44.636
Target Compounds							
2)	A alpha-BHC	4.001	3.393	339.7E6	1483.9E6	58.278	55.376
3)	MA gamma-BHC...	4.332	3.729	319.1E6	1357.0E6	57.165	54.815
4)	MA Heptachlor	4.931	4.082	282.5E6	1246.7E6	51.812	49.793
5)	MB Aldrin	5.273	4.368	301.2E6	1322.1E6	56.698	54.595
6)	B beta-BHC	4.518	4.025	121.5E6	572.2E6	55.031	53.248
7)	B delta-BHC	4.766	4.261	307.0E6	1351.5E6	59.807	54.336
8)	B Heptachlo...	5.692	4.871	258.7E6	1193.8E6	53.631	54.515
9)	A Endosulfan I	6.076	5.246	241.9E6	1103.0E6	53.657	52.923
10)	B gamma-Chl...	5.947	5.125	271.0E6	1303.4E6	56.596	54.485
11)	B alpha-Chl...	6.028	5.189	264.9E6	1220.7E6	54.740	53.314
12)	B 4,4'-DDE	6.197	5.374	230.1E6	1202.0E6	52.764	52.451
13)	MA Dieldrin	6.348	5.512	256.7E6	1221.7E6	53.510	52.641
14)	MA Endrin	6.575	5.788	205.2E6	1082.4E6	49.739	49.849
15)	B Endosulfa...	6.787	6.079	208.1E6	1036.8E6	51.560	51.334
16)	A 4,4'-DDD	6.706	5.928	181.6E6	970.4E6	52.921	50.729
17)	MA 4,4'-DDT	7.022	6.182	158.6E6	868.9E6	41.870	42.845
18)	B Endrin al...	6.916	6.257	146.4E6	744.7E6	47.779	48.636
19)	B Endosulfa...	7.151	6.481	187.2E6	956.9E6	48.781	48.769
20)	A Methoxychlor	7.495	6.753	84285924	469.6E6	41.713	43.610
21)	B Endrin ke...	7.631	6.990	198.4E6	1021.1E6	48.704	47.268
22)	Mirex	8.115	7.184	140.5E6	762.7E6	46.229	45.733

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062725\
Data File : PD089188.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2025 12:53
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 05:37:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

