

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062918\
 Data File : PD048253.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Jun 2018 11:51
 Operator : UA/AJ
 Sample : J3704-12
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 22:59:45 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_D\Method\PD062818CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 29 16:27:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
Target Compounds						
3) MA gamma-BHC...	4.106	4.672	88621136	1320.5E6	47.260	45.522
4) MA Heptachlor	4.394	5.187	86830242	1309.0E6	49.078	45.245
5) MB Aldrin	4.636	5.492	82498449	1240.9E6	47.939	44.580
13) MA Dieldrin	5.647	6.483	171.5E6	2190.4E6	99.180	90.768
14) MA Endrin	5.903	6.699	157.4E6	1879.8E6	104.552	95.182
17) MA 4,4'-DDT	6.268	7.115	144.8E6	1781.4E6	100.007	92.914

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062918\
Data File : PD048253.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Jun 2018 11:51
Operator : UA/AJ
Sample : J3704-12
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
PEST-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 22:59:45 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_D\Method\PD062818CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 29 16:27:50 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

