

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062921\
 Data File : PD064035.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2021 23:41
 Operator : AR\AJ
 Sample : M2808-20MS
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BGD3MS

Manual Integrations
 APPROVED

mohammad
 7/1/2021 12:01:32 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 04:07:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061421CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 21 04:24:00 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.434	3.969	12339480	109.7E6	7.938	6.925
27)	SA Decachlor...	8.183	9.009	35505429	189.3E6	20.512	13.649m#
Target Compounds							
2)	A alpha-BHC	3.868	4.356	82202150	837.0E6	35.143m	33.688m
3)	MA gamma-BHC...	4.153	4.642	89938857	921.5E6	39.520	40.384
4)	MA Heptachlor	4.452	5.147	78242724	785.5E6	35.838	33.719m
5)	MB Aldrin	4.699	5.453	71880887	807.6E6	35.465	36.660
6)	B beta-BHC	4.404	4.815	36287846	284.5E6	37.372	29.174m
7)	B delta-BHC	4.607	5.029	34358961	373.1E6	15.918	15.924
8)	B Heptachlo...	5.141	5.838	87431596	533.6E6	46.107m	25.937m#
9)	A Endosulfan I	5.482	6.195	48530604	586.7E6	26.522	29.613
10)	B trans-Chl...	5.368	6.071	62797170	928.0E6	32.292m	44.117 #
11)	B cis-Chlor...	5.428	6.143	60078143	726.0E6	31.621	36.046m
12)	B 4,4'-DDE	5.591	6.300	135.6E6	1014.2E6	68.703	50.196 #
13)	MA Dieldrin	5.724	6.453	160.2E6	1104.6E6	76.530m	52.804 #
14)	MA Endrin	5.982	6.670	130.6E6	957.9E6	72.214	54.746
15)	B Endosulfa...	6.255	6.878	114.1E6	1037.6E6	69.831	64.127
16)	A 4,4'-DDD	6.108	6.789	88678778	976.5E6	53.711	58.012m
17)	MA 4,4'-DDT	6.349	7.090	97484739	838.6E6	57.923	50.940
18)	B Endrin al...	6.424	7.000	86268554	850.4E6	60.426	60.452m
19)	B Endosulfa...	6.639	7.225	89527889	1100.5E6	57.671	66.868
20)	A Methoxychlor	6.895	7.548	298.9E6	3426.3E6	324.157	442.863 #
21)	B Endrin ke...	7.129	7.696	95209074	1256.5E6	50.187	79.179 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062921\
Data File : PD064035.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2021 23:41
Operator : AR\AJ
Sample : M2808-20MS
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
BGDX3MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 04:07:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061421CLP.M
Quant Title : GC Extractables
QLast Update : Mon Jun 21 04:24:00 2021
Response via : Initial Calibration
Integrator: ChemStation

Manual Integrations
APPROVED

mohammad
7/1/2021 12:01:32 PM

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

