

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062923\
 Data File : PD076224.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2023 15:24
 Operator : AR\AJ
 Sample : 03471-03
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 VNJ-232

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/30/2023
 Supervised By : Ankita Jodhani 06/30/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 20:34:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 16 10:43:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.767	44351657	22768717	20.821	20.986
28)	SA Decachlor...	9.087	7.911	45912466	20190157	18.123	20.649
Target Compounds							
8)	B Heptachlo...	5.695	4.720	4633375	1877799	1.438m	1.367m
10)	B gamma-Chl...	5.953	4.971	7414382	1791956	2.302m	1.306m#
11)	B alpha-Chl...	6.037	5.035	17787677	3303295	5.502m	2.408m#
12)	B 4,4'-DDE	6.204	5.224	2528404	2663784	0.885	2.387m#
17)	MA 4,4'-DDT	7.040	6.030	5737695	3658392	2.528m	4.504 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062923\
Data File : PD076224.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2023 15:24
Operator : AR\AJ
Sample : 03471-03
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
VNJ-232

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/30/2023
Supervised By : Ankita Jodhani 06/30/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 20:34:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061523.M
Quant Title : GC Extractables
QLast Update : Fri Jun 16 10:43:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

