

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD063022\
 Data File : PD070629.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jun 2022 22:41
 Operator : AR\AJ
 Sample : INDB101
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB1031

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 03:14:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD063022CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 01 03:07:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.487	4.126	11238264	120.5E6	5.238	5.509
27)	SA Decachlor...	8.267	9.319	23538512	76004339	11.083	12.293
Target Compounds							
5)	MB Aldrin	4.766	5.651	15240503	134.9E6	4.942	5.464
6)	B beta-BHC	4.464	5.000	8365586	92419063	5.862	6.623
7)	B delta-BHC	4.670	5.223	15103163	147.4E6	4.878	5.385
8)	B Heptachlo...	5.211	6.046	14736046	119.2E6	5.192	5.688
10)	B trans-Chl...	5.439	6.283	15245481	109.9E6	5.133	5.510
11)	B cis-Chlor...	5.497	6.356	14969029	107.7E6	5.201	5.594
12)	B 4,4'-DDE	5.664	6.510	26981846	187.6E6	9.647	10.760
15)	B Endosulfa...	6.329	7.109	24188963	140.6E6	10.348	11.216
18)	B Endrin al...	6.499	7.234	20130355	113.2E6	10.508	11.037
19)	B Endosulfa...	6.712	7.459	23670150	127.6E6	10.365	11.032
21)	B Endrin ke...	7.206	7.941	25169230	120.8E6	10.264	11.212

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD063022\
Data File : PD070629.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jun 2022 22:41
Operator : AR\AJ
Sample : INDB101
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 03:14:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD063022CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Fri Jul 01 03:07:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

