

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD D\Data\PD070318\
 Data File : PD048296.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Jul 2018 23:31
 Operator : UA/AJ
 Sample : J3721-14
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 F1H00

Manual Integrations
 APPROVED

Sohil
 7/9/2018 4:28:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 15:10:32 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_D\Method\PD062818CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 29 16:27:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.393	4.005	20837201	202.3E6	17.244m	9.939m#
27)	SA Decachlor...	8.081	9.033	957.9E6	17917.4E6	716.549m	1011.135m#
Target Compounds							
12)	B 4,4'-DDE	5.523	6.332	337.7E6	1089.9E6	202.449	45.601 #
13)	MA Dieldrin	5.651	6.495	392.5E6	4335.8E6	226.986m	179.672m
14)	MA Endrin	5.898	6.705	380.2E6	2532.9E6	252.520	128.253m#
16)	A 4,4'-DDD	6.032	6.805	193.7E6	4632.9E6	139.917m	246.836 #
17)	MA 4,4'-DDT	6.266	7.116	636.7E6	4771.5E6	439.791	248.869 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Data\PD070318\
Data File : PD048296.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Jul 2018 23:31
Operator : UA/AJ
Sample : J3721-14
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
F1H00

Manual Integrations
APPROVED

Sohil
7/9/2018 4:28:23 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 15:10:32 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_D\Method\PD062818CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 29 16:27:50 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




File :Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070318\PD048296.D
Operator : UA/AJ
Acquired : 03 Jul 2018 23:31 using AcqMethod 8081C.M
Instrument : ECD7
Sample Name: J3721-14
Misc Info :
Vial Number: 31

Instrument :
ECD_D
ClientSampleId :
F1H00

Manual Integrations
APPROVED

Sohil
7/9/2018 4:28:23 PM



File :Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070318\PD048296.D
Operator : UA/AJ
Acquired : 03 Jul 2018 23:31 using AcqMethod 8081C.M
Instrument : ECD7
Sample Name: J3721-14
Misc Info :
Vial Number: 31

Instrument :
ECD_D
ClientSampled :
F1H00

Manual Integrations
APPROVED

Sohil
7/9/2018 4:28:23 PM

