

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0070718.M
 Title : GC EXTRACTABLES
 Last Update : Mon Jul 09 07:25:43 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0046451.D 1000 =P0046449.D 750 =P0046450.D
 250 =P0046452.D 50 =P0046453.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.351	2.293	2.343	2.500	2.463	2.390	E5	3.65
2) SA	Decachlorobiphenyl	1.879	1.716	1.800	2.034	2.099	1.906	E5	8.36
3) L1	AR-1016-1	5.462	5.258	5.332	5.979	6.025	5.611	E3	6.50
4) L1	AR-1016-2	7.034	6.627	6.884	7.543	7.605	7.139	E3	5.94
5) L1	AR-1016-3	5.970	5.624	5.816	6.350	6.267	6.005	E3	5.06
6) L1	AR-1016-4	5.715	5.388	5.578	6.151	6.306	5.828	E3	6.66
7) L1	AR-1016-5	6.320	5.992	6.159	6.709	7.091	6.454	E3	6.88
8) L2	AR-1221-1	3.183					3.183	E3	0.00
9) L2	AR-1221-2	2.365					2.365	E3	0.00
10) L2	AR-1221-3	7.559					7.559	E3	0.00
11) L3	AR-1232-1	2.444					2.444	E3	0.00
12) L3	AR-1232-2	3.540					3.540	E3	0.00
13) L3	AR-1232-3	5.076					5.076	E3	0.00
14) L3	AR-1232-4	3.259					3.259	E3	0.00
15) L3	AR-1232-5	2.655					2.655	E3	0.00
16) L4	AR-1242-1	1.017					1.017	E4	0.00
17) L4	AR-1242-2	6.474					6.474	E3	0.00
18) L4	AR-1242-3	5.448					5.448	E3	0.00
19) L4	AR-1242-4	5.342					5.342	E3	0.00
20) L4	AR-1242-5	6.032					6.032	E3	0.00
21) L5	AR-1248-1	7.934					7.934	E3	0.00
22) L5	AR-1248-2	7.081					7.081	E3	0.00
23) L5	AR-1248-3	9.162					9.162	E3	0.00
24) L5	AR-1248-4	8.870					8.870	E3	0.00
25) L5	AR-1248-5	9.101					9.101	E3	0.00
26) L6	AR-1254-1	1.394					1.394	E4	0.00
27) L6	AR-1254-2	8.808					8.808	E3	0.00
28) L6	AR-1254-3	1.529					1.529	E4	0.00
29) L6	AR-1254-4	1.148					1.148	E4	0.00
30) L6	AR-1254-5	8.875					8.875	E3	0.00
31) L7	AR-1260-1	1.133	1.023	1.069	1.207	1.303	1.147	E4	9.71
32) L7	AR-1260-2	1.318	1.207	1.261	1.427	1.535	1.350	E4	9.77
33) L7	AR-1260-3	1.574	1.446	1.508	1.688	1.593	1.562	E4	5.85
34) L7	AR-1260-4	2.065	1.929	2.004	2.207	2.304	2.102	E4	7.24
35) L7	AR-1260-5	1.315	1.218	1.275	1.420	1.509	1.347	E4	8.65
36) L8	AR-1262-1	9.526					9.526	E3	0.00
37) L8	AR-1262-2	1.114					1.114	E4	0.00
38) L8	AR-1262-3	1.432					1.432	E4	0.00
39) L8	AR-1262-4	1.345					1.345	E4	0.00
40) L8	AR-1262-5	2.420					2.420	E4	0.00
41) L9	AR-1268-1	2.438					2.438	E4	0.00
42) L9	AR-1268-2	2.266					2.266	E4	0.00
43) L9	AR-1268-3	1.889					1.889	E4	0.00
44) L9	AR-1268-4	8.481					8.481	E3	0.00
45) L9	AR-1268-5	5.531					5.531	E4	0.00

Signal #2 Calibration Files

500 =P0046451.D 1000 =P0046449.D 750 =P0046450.D
 250 =P0046452.D 50 =P0046453.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.136	2.051	2.097	2.214	2.123	2.124	E5	2.81
2) SA	Decachlorobiphenyl	1.923	1.804	1.878	2.035	2.060	1.940	E5	5.53

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0070718.M
 Title : GC EXTRACTABLES
 Last Update : Mon Jul 09 07:25:43 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0046451.D 1000 =P0046449.D 750 =P0046450.D
 250 =P0046452.D 50 =P0046453.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	5.544	5.114	5.311	5.737	5.641	5.469	E3	4.64
4)	L1	AR-1016-2	4.692	4.274	4.444	5.068	4.648	4.625	E3	6.46
5)	L1	AR-1016-3	5.759	5.287	5.484	6.086	4.694	5.462	E3	9.60
6)	L1	AR-1016-4	6.346	5.831	6.019	6.673	6.160	6.206	E3	5.19
7)	L1	AR-1016-5	5.533	5.113	5.311	5.760	5.322	5.408	E3	4.56
8)	L2	AR-1221-1	2.674					2.674	E3	0.00
9)	L2	AR-1221-2	2.156					2.156	E3	0.00
10)	L2	AR-1221-3	6.851					6.851	E3	0.00
11)	L3	AR-1232-1	2.321					2.321	E3	0.00
12)	L3	AR-1232-2	3.445					3.445	E3	0.00
13)	L3	AR-1232-3	4.771					4.771	E3	0.00
14)	L3	AR-1232-4	3.514					3.514	E3	0.00
15)	L3	AR-1232-5	2.785					2.785	E3	0.00
16)	L4	AR-1242-1	9.580					9.580	E3	0.00
17)	L4	AR-1242-2	5.550					5.550	E3	0.00
18)	L4	AR-1242-3	5.193					5.193	E3	0.00
19)	L4	AR-1242-4	5.675					5.675	E3	0.00
20)	L4	AR-1242-5	5.180					5.180	E3	0.00
21)	L5	AR-1248-1	8.456					8.456	E3	0.00
22)	L5	AR-1248-2	6.383					6.383	E3	0.00
23)	L5	AR-1248-3	1.231					1.231	E4	0.00
24)	L5	AR-1248-4	8.705					8.705	E3	0.00
25)	L5	AR-1248-5	5.855					5.855	E3	0.00
26)	L6	AR-1254-1	1.337					1.337	E4	0.00
27)	L6	AR-1254-2	1.156					1.156	E4	0.00
28)	L6	AR-1254-3	1.968					1.968	E4	0.00
29)	L6	AR-1254-4	1.261					1.261	E4	0.00
30)	L6	AR-1254-5	9.070					9.070	E3	0.00
31)	L7	AR-1260-1	1.291	1.181	1.229	1.374	1.443	1.303	E4	8.14
32)	L7	AR-1260-2	1.570	1.440	1.496	1.668	1.797	1.594	E4	8.90
33)	L7	AR-1260-3	1.752	1.615	1.682	1.859	2.052	1.792	E4	9.55
34)	L7	AR-1260-4	2.754	2.564	2.654	2.909	2.974	2.771	E4	6.17
35)	L7	AR-1260-5	1.870	1.761	1.825	1.988	2.132	1.915	E4	7.67
36)	L8	AR-1262-1	1.310					1.310	E4	0.00
37)	L8	AR-1262-2	1.295					1.295	E4	0.00
38)	L8	AR-1262-3	1.751					1.751	E4	0.00
39)	L8	AR-1262-4	1.529					1.529	E4	0.00
40)	L8	AR-1262-5	3.087					3.087	E4	0.00
41)	L9	AR-1268-1	2.798					2.798	E4	0.00
42)	L9	AR-1268-2	2.755					2.755	E4	0.00
43)	L9	AR-1268-3	2.232					2.232	E4	0.00
44)	L9	AR-1268-4	9.518					9.518	E3	0.00
45)	L9	AR-1268-5	6.313					6.313	E4	0.00

(#) = Out of Range