

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070523\
 Data File : PD076337.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2023 11:09
 Operator : AR\AJ
 Sample : INDA301
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 05 14:08:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070523CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 05 14:07:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.766	43077995	20388665	20.000	20.000
27)	SA Decachlor...	9.086	7.908	100.2E6	35916538	40.000	40.000
Target Compounds							
2)	A alpha-BHC	4.002	3.267	71499255	30166227	20.000	20.000
3)	MA gamma-BHC...	4.335	3.597	67013205	28279048	20.000	20.000
4)	MA Heptachlor	4.926	3.937	69886124	27965123	20.000	20.000
9)	A Endosulfan I	6.084	5.090	57000402	22732359	20.000	20.000
13)	MA Dieldrin	6.359	5.354	126.6E6	51889624	40.000	40.000
14)	MA Endrin	6.590	5.630	101.8E6	44856402	40.000	40.000
16)	A 4,4'-DDD	6.723	5.777	89860980	32160152	40.000	40.000
17)	MA 4,4'-DDT	7.038	6.029	89811741	33394002	40.000	40.000
20)	A Methoxychlor	7.514	6.604	265.1E6	92215994	200.000	200.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070523\
Data File : PD076337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2023 11:09
Operator : AR\AJ
Sample : INDA301
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 05 14:08:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070523CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jul 05 14:07:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

