

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070720\
 Data File : PD058478.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2020 13:43
 Operator : AJ\MA
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD070720

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 13:57:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070720.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 07 12:46:46 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.209	3.866	90317888	122.9E6	49.535	49.727
2)	SA Decachlor...	7.845	8.859	86463874	101.5E6	47.880	46.553
Target Compounds							
2)	A alpha-BHC	3.631	4.248	145.8E6	193.6E6	50.919	51.358
3)	MA gamma-BHC...	3.902	4.529	128.9E6	172.9E6	50.335	50.005
4)	MA Heptachlor	4.182	5.028	92736528	155.5E6	51.145	49.390
5)	MB Aldrin	4.417	5.329	118.9E6	156.6E6	49.903	49.368
6)	B beta-BHC	4.151	4.700	49568515	68661124	48.380	48.476
7)	B delta-BHC	4.342	4.912	127.6E6	159.8E6	50.309	51.967
8)	B Heptachlo...	4.849	5.712	100.4E6	140.3E6	51.145	48.196
9)	A Endosulfan I	5.176	6.065	99262047	129.3E6	50.093	49.032
10)	B gamma-Chl...	5.069	5.945	110.6E6	143.5E6	49.675	49.302
11)	B alpha-Chl...	5.125	6.017	109.5E6	140.4E6	49.443	48.816
12)	B 4,4'-DDE	5.294	6.177	108.1E6	137.1E6	50.104	49.866
13)	MA Dieldrin	5.414	6.320	95215793	140.8E6	53.394	49.296
14)	MA Endrin	5.665	6.535	68184587	96159952	47.877	45.180
15)	B Endosulfa...	5.938	6.745	90570785	117.0E6	50.193	47.400
16)	A 4,4'-DDD	5.802	6.664	69208204	113.1E6	52.909	49.045
17)	MA 4,4'-DDT	6.036	6.961	73332567	109.5E6	49.753	47.710
18)	B Endrin al...	6.106	6.868	74016565	103.4E6	53.143	48.350
19)	B Endosulfa...	6.315	7.091	78415651	108.8E6	49.279	47.466
20)	A Methoxychlor	6.584	7.421	29628117	55098732	50.291	45.833
21)	B Endrin ke...	6.796	7.557	100.6E6	133.6E6	55.320	49.076

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070720\
Data File : PD058478.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2020 13:43
Operator : AJ\MA
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD070720

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 13:57:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070720.M
Quant Title : GC Extractables
QLast Update : Tue Jul 07 12:46:46 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

