

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070822\
 Data File : PD070767.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2022 15:08
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 16:17:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070522.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 07 15:20:45 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.485	4.124	115.0E6	1280.8E6	47.105	49.526
28)	SA Decachlor...	8.262	9.313	118.0E6	361.8E6	47.575	43.982
Target Compounds							
2)	A alpha-BHC	3.928	4.528	169.5E6	1883.9E6	48.282	51.264
3)	MA gamma-BHC...	4.213	4.821	159.8E6	1660.3E6	46.971	51.102
4)	MA Heptachlor	4.514	5.337	160.7E6	1632.8E6	47.641	52.708
5)	MB Aldrin	4.763	5.648	161.4E6	1480.0E6	48.879	52.496
6)	B beta-BHC	4.462	4.997	66440817	736.1E6	45.529	53.327
7)	B delta-BHC	4.668	5.220	159.5E6	1591.6E6	48.749	55.639
8)	B Heptachlo...	5.209	6.043	143.1E6	1216.7E6	47.494	53.470
9)	A Endosulfan I	5.549	6.408	129.7E6	1067.5E6	46.336	52.126
10)	B gamma-Chl...	5.436	6.279	148.3E6	1224.8E6	47.818	53.710
11)	B alpha-Chl...	5.495	6.352	144.7E6	1153.5E6	47.577	52.363
12)	B 4,4'-DDE	5.660	6.507	140.9E6	1065.3E6	48.463	52.097
13)	MA Dieldrin	5.794	6.668	145.0E6	1055.6E6	48.199	52.082
14)	MA Endrin	6.052	6.892	127.0E6	737.9E6	47.230	47.933
15)	B Endosulfa...	6.326	7.105	118.6E6	739.7E6	49.938	50.384
16)	A 4,4'-DDD	6.178	7.007	119.8E6	750.7E6	49.106	53.044
17)	MA 4,4'-DDT	6.421	7.308	123.2E6	756.1E6	49.352	49.626
18)	B Endrin al...	6.496	7.231	91044239	580.6E6	48.974	50.704
19)	B Endosulfa...	6.709	7.455	117.2E6	698.8E6	48.257	50.925
20)	A Methoxychlor	6.972	7.767	67262830	335.4E6	47.059	48.910
21)	B Endrin ke...	7.204	7.936	136.0E6	638.8E6	51.107	49.783
22)	Mirex	7.391	8.390	103.6E6	388.2E6	47.322	48.608

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070822\
Data File : PD070767.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2022 15:08
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 16:17:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070522.M
Quant Title : GC Extractables
QLast Update : Thu Jul 07 15:20:45 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

