

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070920\
 Data File : PD058507.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2020 12:54
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 01:19:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070720.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 08 08:01:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.210	3.866	75758887	104.8E6	41.550	42.428
2)	SA Decachlor...	7.845	8.858	75199705	91450140	41.642	41.944
Target Compounds							
2)	A alpha-BHC	3.632	4.249	124.7E6	165.1E6	43.550	43.796
3)	MA gamma-BHC...	3.903	4.529	112.7E6	150.7E6	44.027	43.589
4)	MA Heptachlor	4.182	5.029	85208568	138.9E6	46.993	44.106
5)	MB Aldrin	4.417	5.329	100.7E6	136.1E6	42.272	42.905
6)	B beta-BHC	4.152	4.700	43114144	60144647	42.080	42.464
7)	B delta-BHC	4.343	4.912	112.0E6	130.1E6	44.171	42.311
8)	B Heptachlo...	4.849	5.713	88277704	125.2E6	44.982	43.007
9)	A Endosulfan I	5.176	6.065	85050192	115.1E6	42.921	43.646
10)	B gamma-Chl...	5.069	5.945	94605255	126.8E6	42.493	43.573
11)	B alpha-Chl...	5.125	6.017	93521952	124.9E6	42.236	43.441
12)	B 4,4'-DDE	5.295	6.177	91527483	121.0E6	42.410	44.026
13)	MA Dieldrin	5.413	6.320	84842443	124.4E6	47.577	43.557
14)	MA Endrin	5.665	6.536	72997219	104.1E6	51.256	48.915
15)	B Endosulfa...	5.938	6.745	79539520	106.7E6	44.080	43.236
16)	A 4,4'-DDD	5.802	6.664	61686736	100.2E6	47.159	43.442
17)	MA 4,4'-DDT	6.036	6.961	67744427	102.5E6	45.962	44.686
18)	B Endrin al...	6.106	6.869	63511994	90314703	45.601	42.233
19)	B Endosulfa...	6.316	7.092	72423934	101.0E6	45.513	44.037
20)	A Methoxychlor	6.584	7.421	29434590	54951649	49.962	45.711
21)	B Endrin ke...	6.797	7.557	86546299	114.7E6	47.043	42.150
22)	Mirex	6.966	8.007	64111051	82389488	42.563	42.703

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD070920\
Data File : PD058507.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2020 12:54
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:19:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070720.M
Quant Title : GC Extractables
QLast Update : Wed Jul 08 08:01:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

