

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071025\
 Data File : PD089429.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2025 11:44
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 22:56:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.556	2.878	146.2E6	911.2E6	51.240	53.749
28)	SA Decachlor...	9.082	8.072	198.2E6	1067.2E6	50.452	53.983
Target Compounds							
2)	A alpha-BHC	4.006	3.391	302.9E6	1406.2E6	51.965	52.476
3)	MA gamma-BHC...	4.337	3.727	287.1E6	1283.8E6	51.436	51.854
4)	MA Heptachlor	4.937	4.081	282.0E6	1271.8E6	51.716	50.792
5)	MB Aldrin	5.278	4.367	270.8E6	1277.6E6	50.968	52.758
6)	B beta-BHC	4.523	4.024	110.0E6	552.8E6	49.826	51.441
7)	B delta-BHC	4.772	4.260	276.5E6	1304.4E6	53.871	52.439
8)	B Heptachlo...	5.698	4.871	240.9E6	1183.7E6	49.938	54.055
9)	A Endosulfan I	6.082	5.246	229.3E6	1103.9E6	50.874	52.963
10)	B gamma-Chl...	5.953	5.125	244.8E6	1248.4E6	51.134	52.183
11)	B alpha-Chl...	6.035	5.189	245.4E6	1191.7E6	50.710	52.045
12)	B 4,4'-DDE	6.203	5.374	221.5E6	1203.7E6	50.796	52.526
13)	MA Dieldrin	6.355	5.512	245.2E6	1212.7E6	51.110	52.251
14)	MA Endrin	6.582	5.788	199.0E6	1056.9E6	48.237	48.674
15)	B Endosulfa...	6.794	6.080	205.9E6	1058.5E6	51.019	52.409
16)	A 4,4'-DDD	6.713	5.929	176.1E6	996.4E6	51.300	52.087
17)	MA 4,4'-DDT	7.029	6.183	190.5E6	1051.5E6	50.290	51.849
18)	B Endrin al...	6.923	6.258	148.3E6	767.9E6	48.404	50.151
19)	B Endosulfa...	7.157	6.482	190.8E6	1023.9E6	49.713	52.182
20)	A Methoxychlor	7.501	6.754	96213416	521.4E6	47.616	48.419
21)	B Endrin ke...	7.638	6.991	206.1E6	1136.3E6	50.595	52.604
22)	Mirex	8.121	7.184	150.7E6	865.6E6	49.564	51.898

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071025\
Data File : PD089429.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 11:44
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:56:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

